

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100224\
 Data File : VY019763.D
 Acq On : 02 Oct 2024 14:36
 Operator : SY/MD
 Sample : P4236-02
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP4-VOC

Quant Time: Oct 03 02:02:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	1785	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	3572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1613	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	556	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	1815	109.809	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	219.620%#
35) Dibromofluoromethane	7.634	113	1332	65.769	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	131.540%
50) Toluene-d8	10.115	98	3333	44.494	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.980%
62) 4-Bromofluorobenzene	12.414	95	399	13.853	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	27.700%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100224\
Data File : VY019763.D
Acq On : 02 Oct 2024 14:36
Operator : SY/MD
Sample : P4236-02
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP4-VOC

Quant Time: Oct 03 02:02:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

