

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100224\
 Data File : VY019766.D
 Acq On : 02 Oct 2024 15:47
 Operator : SY/MD
 Sample : P4242-03
 Misc : 4.16g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-Q-VOC

Quant Time: Oct 03 02:03:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	69308	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	114158	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	82043	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	23211	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	28748	44.794	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.580%
35) Dibromofluoromethane	7.634	113	33132	51.188	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.380%
50) Toluene-d8	10.103	98	125315	52.344	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.680%
62) 4-Bromofluorobenzene	12.401	95	31675	34.410	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	68.820%
Target Compounds						
78) n-propylbenzene	12.590	91	3178	1.592	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100224\
Data File : VY019766.D
Acq On : 02 Oct 2024 15:47
Operator : SY/MD
Sample : P4242-03
Misc : 4.16g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-Q-VOC

Quant Time: Oct 03 02:03:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

