

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100224\
 Data File : VY019772.D
 Acq On : 02 Oct 2024 18:06
 Operator : SY/MD
 Sample : P4254-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-147-SLUDGE

Quant Time: Oct 03 04:47:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	3124	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	5461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	2788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	410	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	2368	81.860	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	163.720%#
35) Dibromofluoromethane	7.640	113	1912	61.751	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	123.500%
50) Toluene-d8	10.109	98	5472	47.780	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.560%
62) 4-Bromofluorobenzene	12.402	95	481	10.923	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	21.840%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100224\
Data File : VY019772.D
Acq On : 02 Oct 2024 18:06
Operator : SY/MD
Sample : P4254-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-147-SLUDGE

Quant Time: Oct 03 04:47:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

