

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100224\
 Data File : VY019773.D
 Acq On : 02 Oct 2024 18:29
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Oct 03 02:08:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	269385	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	521382	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	460782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	174782	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	152764	61.242	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.480%	
35) Dibromofluoromethane	7.634	113	163436	55.287	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.580%	
50) Toluene-d8	10.103	98	624523	57.117	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 114.240%	
62) 4-Bromofluorobenzene	12.402	95	202929	48.268	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 96.540%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100224\
Data File : VY019773.D
Acq On : 02 Oct 2024 18:29
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Oct 03 02:08:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

