

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100422\
 Data File : VY010785.D
 Acq On : 04 Oct 2022 11:23
 Operator : KP/MD
 Sample : N4936-01RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SASP2-T-2-(0-6)RE

Quant Time: Oct 05 07:25:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	4492	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	6765	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	5584	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	1250	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	1979	43.464	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	86.920%
35) Dibromofluoromethane	7.728	113	1008	22.447	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	44.900%#
50) Toluene-d8	10.179	98	5619	36.588	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	73.180%
62) 4-Bromofluorobenzene	12.483	95	1537	23.680	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	47.360%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100422\
Data File : VY010785.D
Acq On : 04 Oct 2022 11:23
Operator : KP/MD
Sample : N4936-01RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SASP2-T-2-(0-6)RE

Quant Time: Oct 05 07:25:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

