

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100422\
 Data File : VY010798.D
 Acq On : 04 Oct 2022 16:18
 Operator : KP/MD
 Sample : N4943-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-001-0930

Quant Time: Oct 05 07:31:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	5080	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	8249	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	7296	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1919	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	2902	56.358	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.720%
35) Dibromofluoromethane	7.722	113	1248	22.791	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	45.580%#
50) Toluene-d8	10.179	98	6782	36.216	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	72.440%
62) 4-Bromofluorobenzene	12.483	95	2164	27.956	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	55.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100422\
Data File : VY010798.D
Acq On : 04 Oct 2022 16:18
Operator : KP/MD
Sample : N4943-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-001-0930

Quant Time: Oct 05 07:31:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

