



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100424\
Data File : VY019791.D
Acq On : 04 Oct 2024 08:42
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Title : SW846 8260
Last Update : Mon Sep 09 16:00:30 2024



AutoFind: Scans 1751, 1752, 1753; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	35059	PASS
75	95	30	60	55.3	93715	PASS
95	95	100	100	100.0	169493	PASS
96	95	5	9	6.4	10880	PASS
173	174	0.00	2	1.0	1238	PASS
174	95	50	100	70.6	119725	PASS
175	174	5	9	7.5	8994	PASS
176	174	95	101	96.5	115491	PASS
177	176	5	9	6.7	7682	PASS