

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100824\
 Data File : VY019821.D
 Acq On : 08 Oct 2024 14:35
 Operator : SY/MD
 Sample : P4343-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-03-100724

Quant Time: Oct 09 02:07:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	151189	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	276244	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	221826	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	71585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	80783	57.703	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.400%
35) Dibromofluoromethane	7.634	113	86043	54.936	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.880%
50) Toluene-d8	10.109	98	320814	55.378	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.760%
62) 4-Bromofluorobenzene	12.407	95	89640	40.242	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	80.480%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100824\
Data File : VY019821.D
Acq On : 08 Oct 2024 14:35
Operator : SY/MD
Sample : P4343-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-03-100724

Quant Time: Oct 09 02:07:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

