

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100825\
 Data File : VY023398.D
 Acq On : 08 Oct 2025 11:24
 Operator : SY/MD
 Sample : Q3303-01
 Misc : 5.20g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HL-WEST

Quant Time: Oct 09 03:10:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	127568	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	196292	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	158506	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	54504	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	46927	43.991	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.980%
35) Dibromofluoromethane	7.634	113	67132	55.659	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.320%
50) Toluene-d8	10.109	98	212872	47.392	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.780%
62) 4-Bromofluorobenzene	12.407	95	53299	35.942	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	71.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100825\
Data File : VY023398.D
Acq On : 08 Oct 2025 11:24
Operator : SY/MD
Sample : Q3303-01
Misc : 5.20g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HL-WEST

Quant Time: Oct 09 03:10:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

