

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100924\
 Data File : VY019839.D
 Acq On : 09 Oct 2024 17:51
 Operator : SY/MD
 Sample : P4353-02
 Misc : 3.21g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CF-400-VOC-43

Quant Time: Oct 10 05:41:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 10 05:30:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	334359	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	665105	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	588854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	209283	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	211622	51.403	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.800%
35) Dibromofluoromethane	7.634	113	212759	48.476	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.960%
50) Toluene-d8	10.109	98	800868	49.412	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.820%
62) 4-Bromofluorobenzene	12.402	95	249741	42.644	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	85.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100924\
Data File : VY019839.D
Acq On : 09 Oct 2024 17:51
Operator : SY/MD
Sample : P4353-02
Misc : 3.21g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CF-400-VOC-43

Quant Time: Oct 10 05:41:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 10 05:30:07 2024
Response via : Initial Calibration

