

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100925\
 Data File : VY023418.D
 Acq On : 09 Oct 2025 13:21
 Operator : SY/MD
 Sample : Q3310-03
 Misc : 8.17g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SS-01-2-3

Quant Time: Oct 10 04:26:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	579062	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1100821	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	963674	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	310632	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	289335	59.753	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.500%
35) Dibromofluoromethane	7.640	113	371361	54.902	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.800%
50) Toluene-d8	10.109	98	1309292	51.977	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.960%
62) 4-Bromofluorobenzene	12.408	95	352514	42.388	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	84.780%
Target Compounds						Qvalue
16) Acetone	3.879	43	14388	12.481	ug/l	99
20) Methylene Chloride	4.622	84	23323	3.742	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100925\
Data File : VY023418.D
Acq On : 09 Oct 2025 13:21
Operator : SY/MD
Sample : Q3310-03
Misc : 8.17g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SS-01-2-3

Quant Time: Oct 10 04:26:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

