

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101122\
 Data File : VY010919.D
 Acq On : 11 Oct 2022 15:58
 Operator : KP/MD
 Sample : N5018-03
 Misc : 4.91g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SASP2-T-3-(10-14)

Quant Time: Oct 12 04:46:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	257056	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	444852	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	389938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	177839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	154988	59.483	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 118.960%	
35) Dibromofluoromethane	7.716	113	147670	50.007	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.020%	
50) Toluene-d8	10.179	98	537215	53.196	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.400%	
62) 4-Bromofluorobenzene	12.477	95	177553	44.605	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.220%	
Target Compounds						
20) Methylene Chloride	4.710	84	39154	9.247	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101122\
Data File : VY010919.D
Acq On : 11 Oct 2022 15:58
Operator : KP/MD
Sample : N5018-03
Misc : 4.91g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SASP2-T-3-(10-14)

Quant Time: Oct 12 04:46:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

