

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101124\
 Data File : VY019878.D
 Acq On : 11 Oct 2024 17:45
 Operator : SY/MD
 Sample : P4385-17
 Misc : 6.26g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-9

Quant Time: Oct 14 01:50:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 10 05:30:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	253892	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	530815	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	485576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	181975	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	186485	59.653	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.300%
35) Dibromofluoromethane	7.634	113	175204	50.018	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.040%
50) Toluene-d8	10.109	98	646582	49.986	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.980%
62) 4-Bromofluorobenzene	12.402	95	205702	44.011	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	88.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101124\
Data File : VY019878.D
Acq On : 11 Oct 2024 17:45
Operator : SY/MD
Sample : P4385-17
Misc : 6.26g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-9

Quant Time: Oct 14 01:50:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 10 05:30:07 2024
Response via : Initial Calibration

