

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY101419\
 Data File : VY000308.D
 Acq On : 14 Oct 2019 18:41
 Operator : SY/MD
 Sample : K5358-02
 Misc : 5.00g/5.00ML/MSVOA Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 20190897-AMENDED-COMP-KM-CA

Quant Time: Oct 15 02:19:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y101119S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 14 08:10:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	232934	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.70	114	445787	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	396632	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	182188	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.15	65	162096	65.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	131.94%	
35) Dibromofluoromethane	7.73	113	94282	35.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.12%	
50) Toluene-d8	10.18	98	539097	52.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.24%	
62) 4-Bromofluorobenzene	12.48	95	194266	49.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.86%	
Target Compounds						
11) Tert butyl alcohol	4.97	59	90067	350.269	ug/l	99
16) Acetone	3.96	43	35991	45.231	ug/l	96
17) Carbon Disulfide	4.19	76	337963	45.627	ug/l	100
25) 2-Butanone	7.00	43	11225	10.014	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY101419\
Data File : VY000308.D
Acq On : 14 Oct 2019 18:41
Operator : SY/MD
Sample : K5358-02
Misc : 5.00µ/5.00ML/MSVOA Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
20190897-AMENDED-COMP-KM-CA

Quant Time: Oct 15 02:19:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y101119S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 14 08:10:23 2019
Response via : Initial Calibration

