

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101425\
 Data File : VY023460.D
 Acq On : 14 Oct 2025 14:28
 Operator : SY/MD
 Sample : Q3341-01
 Misc : 5.17g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 C-1

Quant Time: Oct 15 03:29:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	550557	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1073471	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1115279	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	474493	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	291839	63.390	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.780%
35) Dibromofluoromethane	7.640	113	365588	55.426	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.860%
50) Toluene-d8	10.109	98	1302826	53.038	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.080%
62) 4-Bromofluorobenzene	12.408	95	443979	54.746	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	109.500%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.629	84	25301	4.269	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101425\
Data File : VY023460.D
Acq On : 14 Oct 2025 14:28
Operator : SY/MD
Sample : Q3341-01
Misc : 5.17g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
C-1

Quant Time: Oct 15 03:29:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

