

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101425\
 Data File : VY023461.D
 Acq On : 14 Oct 2025 14:52
 Operator : SY/MD
 Sample : Q3341-02
 Misc : 4.85g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 C-2

Quant Time: Oct 15 03:29:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	494390	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	985514	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1038095	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	461613	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	276179	66.804	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.600%
35) Dibromofluoromethane	7.640	113	341792	56.443	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.880%
50) Toluene-d8	10.109	98	1204459	53.410	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.820%
62) 4-Bromofluorobenzene	12.408	95	415605	55.821	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	111.640%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.616	84	26547	4.988	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101425\
Data File : VY023461.D
Acq On : 14 Oct 2025 14:52
Operator : SY/MD
Sample : Q3341-02
Misc : 4.85g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
C-2

Quant Time: Oct 15 03:29:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

