

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101521\
 Data File : VY006403.D
 Acq On : 15 Oct 2021 13:56
 Operator : SY/MD
 Sample : M4242-20
 Misc : 6.85G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T2-4-BOTTOM-GRAB

Quant Time: Oct 18 04:45:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	125103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	205198	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	199154	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	92236	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	86145	64.337	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.680%
35) Dibromofluoromethane	7.722	113	61122	55.753	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.500%
50) Toluene-d8	10.185	98	264059	57.552	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.100%
62) 4-Bromofluorobenzene	12.483	95	85253	54.696	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	109.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101521\
Data File : VY006403.D
Acq On : 15 Oct 2021 13:56
Operator : SY/MD
Sample : M4242-20
Misc : 6.85G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T2-4-BOTTOM-GRAB

Quant Time: Oct 18 04:45:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

