

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101521\
 Data File : VY006404.D
 Acq On : 15 Oct 2021 14:19
 Operator : SY/MD
 Sample : M4243-09
 Misc : 6.29G/5ML/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T3-4-BOTTOM-GRAB

Quant Time: Oct 18 04:45:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	124560	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	206819	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	194326	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	81648	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	86605	64.962	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.920%
35) Dibromofluoromethane	7.728	113	59878	54.191	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.380%
50) Toluene-d8	10.185	98	264597	57.217	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.440%
62) 4-Bromofluorobenzene	12.483	95	81706	52.010	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101521\
Data File : VY006404.D
Acq On : 15 Oct 2021 14:19
Operator : SY/MD
Sample : M4243-09
Misc : 6.29G/5ML/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T3-4-BOTTOM-GRAB

Quant Time: Oct 18 04:45:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

