

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101521\
 Data File : VY006412.D
 Acq On : 15 Oct 2021 17:24
 Operator : SY/MD
 Sample : M4249-10
 Misc : 4.67G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP2-(D)

Quant Time: Oct 18 04:46:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	59466	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	94914	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	89236	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	40785	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	39692	62.364	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 124.720%		
35) Dibromofluoromethane	7.728	113	29393	57.964	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 115.920%		
50) Toluene-d8	10.185	98	121480	57.241	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 114.480%		
62) 4-Bromofluorobenzene	12.483	95	36099	50.071	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 100.140%		
Target Compounds						Qvalue
20) Methylene Chloride	4.729	84	7202	8.657	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101521\
Data File : VY006412.D
Acq On : 15 Oct 2021 17:24
Operator : SY/MD
Sample : M4249-10
Misc : 4.67G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP2-(D)

Quant Time: Oct 18 04:46:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

