

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101521\
 Data File : VY006413.D
 Acq On : 15 Oct 2021 17:48
 Operator : SY/MD
 Sample : M4222-01
 Misc : 4.97G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TEST-A(1)

Quant Time: Oct 18 04:46:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	118127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	195812	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	184694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	83620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	74573	58.983	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.960%
35) Dibromofluoromethane	7.728	113	58239	55.670	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.340%
50) Toluene-d8	10.185	98	249729	57.038	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.080%
62) 4-Bromofluorobenzene	12.483	95	75203	50.561	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.120%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.716	84	15809	9.992	ug/l	88
30) Chloroform	7.514	83	2559	1.135	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101521\
Data File : VY006413.D
Acq On : 15 Oct 2021 17:48
Operator : SY/MD
Sample : M4222-01
Misc : 4.97G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TEST-A(1)

Quant Time: Oct 18 04:46:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

