

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101524\
 Data File : VY019903.D
 Acq On : 15 Oct 2024 15:01
 Operator : SY/MD
 Sample : VY1015SBS05
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1015SBS05

Quant Time: Oct 16 02:50:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 10 05:30:07 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
 Supervised By :Semsettin Yesilyurt 10/16/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	156519	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	291370	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	270646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	138870	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	140626	72.969	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 145.940%			
35) Dibromofluoromethane	7.634	113	125588	65.318	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 130.640%			
50) Toluene-d8	10.103	98	441129	62.128	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 124.260%			
62) 4-Bromofluorobenzene	12.401	95	166034	64.716	ug/l	0.00
Spiked Amount 50.000	Range 29 - 146		Recovery = 129.440%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.867	85	49249	33.800	ug/l	99
3) Chloromethane	2.068	50	60729	32.023	ug/l	99
4) Vinyl Chloride	2.202	62	67816	32.497	ug/l	100
5) Bromomethane	2.592	94	42588	32.242	ug/l	98
6) Chloroethane	2.732	64	46827	33.401	ug/l	95
7) Trichlorofluoromethane	3.056	101	108155	34.830	ug/l	95
8) Diethyl Ether	3.458	74	31012	32.861	ug/l	76
9) 1,1,2-Trichlorotrifluo...	3.818	101	64092	35.484	ug/l	91
10) Methyl Iodide	4.007	142	56398	30.873	ug/l	92
11) Tert butyl alcohol	4.848	59	30220	167.379	ug/l #	86
12) 1,1-Dichloroethene	3.793	96	52967	31.544	ug/l #	83
13) Acrolein	3.659	56	24215	170.996	ug/l	98
14) Allyl chloride	4.385	41	102020	33.674	ug/l #	91
15) Acrylonitrile	5.055	53	75611	175.163	ug/l	96
16) Acetone	3.866	43	92835	185.047	ug/l #	82
17) Carbon Disulfide	4.104	76	115611	25.687	ug/l	100
18) Methyl Acetate	4.385	43	36248	33.541	ug/l #	86
19) Methyl tert-butyl Ether	5.116	73	164234	34.045	ug/l	98
20) Methylene Chloride	4.616	84	74423	39.987	ug/l #	83
21) trans-1,2-Dichloroethene	5.116	96	57423	31.421	ug/l	88
22) Diisopropyl ether	6.018	45	237258	35.953	ug/l #	87
23) Vinyl Acetate	5.964	43	650190	171.863	ug/l #	91
24) 1,1-Dichloroethane	5.915	63	127452	34.744	ug/l	97
25) 2-Butanone	6.896	43	120980	180.159	ug/l #	81
26) 2,2-Dichloropropane	6.884	77	107155	32.800	ug/l	95
27) cis-1,2-Dichloroethene	6.896	96	75084	33.246	ug/l	82
28) Bromochloromethane	7.250	49	53201	32.950	ug/l #	73
29) Tetrahydrofuran	7.268	42	66203	172.663	ug/l #	81
30) Chloroform	7.421	83	132204	35.354	ug/l	97
31) Cyclohexane	7.701	56	102365	31.789	ug/l	87
32) 1,1,1-Trichloroethane	7.616	97	113393	34.601	ug/l	95
36) 1,1-Dichloropropene	7.835	75	84268	30.421	ug/l	95
37) Ethyl Acetate	6.982	43	46268	32.033	ug/l	96
38) Carbon Tetrachloride	7.817	117	92713	31.221	ug/l	95
39) Methylcyclohexane	9.109	83	105675	30.277	ug/l #	86
40) Benzene	8.079	78	263002	31.373	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101524\
 Data File : VY019903.D
 Acq On : 15 Oct 2024 15:01
 Operator : SY/MD
 Sample : VY1015SBS05
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1015SBS05

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
 Supervised By :Semsettin Yesilyurt 10/16/2024

Quant Time: Oct 16 02:50:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 10 05:30:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.219	41	27020	34.564	ug/l #	83
42) 1,2-Dichloroethane	8.158	62	78283	32.504	ug/l	95
43) Isopropyl Acetate	8.195	43	92669	31.459	ug/l #	75
44) Trichloroethene	8.866	130	62756	30.908	ug/l	91
45) 1,2-Dichloropropane	9.140	63	69211	32.985	ug/l	100
46) Dibromomethane	9.231	93	34709	30.276	ug/l	91
47) Bromodichloromethane	9.420	83	98811	32.472	ug/l	97
48) Methyl methacrylate	9.219	41	41144	31.753	ug/l #	81
49) 1,4-Dioxane	9.219	88	8282	645.471	ug/l #	87
51) 4-Methyl-2-Pentanone	9.993	43	244204	162.532	ug/l	88
52) Toluene	10.170	92	166746	31.865	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	83950	30.662	ug/l	98
54) cis-1,3-Dichloropropene	9.853	75	99743	30.874	ug/l #	83
55) 1,1,2-Trichloroethane	10.573	97	48372	32.026	ug/l	96
56) Ethyl methacrylate	10.438	69	67984	31.368	ug/l #	73
57) 1,3-Dichloropropane	10.713	76	85910	32.389	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.713	63	178150	176.572	ug/l	93
59) 2-Hexanone	10.762	43	179795	167.542	ug/l	86
60) Dibromochloromethane	10.908	129	60998	31.464	ug/l	100
61) 1,2-Dibromoethane	11.011	107	43596	31.941	ug/l	95
64) Tetrachloroethene	10.646	164	53292	27.683	ug/l	94
65) Chlorobenzene	11.438	112	180178	29.106	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.511	131	61650	29.436	ug/l	99
67) Ethyl Benzene	11.517	91	328219	29.188	ug/l	99
68) m/p-Xylenes	11.627	106	240397	57.885	ug/l	92
69) o-Xylene	11.950	106	116819	29.276	ug/l	93
70) Styrene	11.969	104	195685	29.067	ug/l	95
71) Bromoform	12.127	173	32188	28.709	ug/l #	95
73) Isopropylbenzene	12.249	105	318461	27.264	ug/l	98
74) N-amyl acetate	12.066	43	80584	26.047	ug/l #	87
75) 1,1,2,2-Tetrachloroethane	12.505	83	58337	28.107	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	38151m	14.024	ug/l	
77) Bromobenzene	12.530	156	66080	26.659	ug/l	84
78) n-propylbenzene	12.590	91	396766	28.111	ug/l	96
79) 2-Chlorotoluene	12.676	91	224393	27.732	ug/l	93
80) 1,3,5-Trimethylbenzene	12.731	105	259309	27.489	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.304	75	16845	24.993	ug/l #	81
82) 4-Chlorotoluene	12.773	91	224020	27.076	ug/l	97
83) tert-Butylbenzene	12.993	119	234509	27.750	ug/l	93
84) 1,2,4-Trimethylbenzene	13.042	105	255465	27.315	ug/l	97
85) sec-Butylbenzene	13.170	105	361040	28.595	ug/l	97
86) p-Isopropyltoluene	13.285	119	285701	27.866	ug/l	96
87) 1,3-Dichlorobenzene	13.285	146	136318	27.233	ug/l	97
88) 1,4-Dichlorobenzene	13.365	146	132600	26.964	ug/l	94
89) n-Butylbenzene	13.615	91	284233	28.391	ug/l	98
90) Hexachloroethane	13.877	117	54255	27.176	ug/l	87
91) 1,2-Dichlorobenzene	13.657	146	119633	27.200	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.273	75	9030	27.261	ug/l	71
93) 1,2,4-Trichlorobenzene	14.919	180	65502	26.559	ug/l	97
94) Hexachlorobutadiene	15.023	225	37465	27.324	ug/l	96
95) Naphthalene	15.139	128	124244	26.710	ug/l	98
96) 1,2,3-Trichlorobenzene	15.328	180	55142	26.450	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101524\
Data File : VY019903.D
Acq On : 15 Oct 2024 15:01
Operator : SY/MD
Sample : VY1015SBS05
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024

Quant Time: Oct 16 02:50:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 10 05:30:07 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY01524\
 Data File : VY019903.D
 Acq On : 15 Oct 2024 15:01
 Operator : SY/MD
 Sample : VY015SBS05
 Misc : 5.00g/5.0mL/MSVOA_Y/soil
 ALS Vial : 9 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VY015SBS05

Quant Time: Oct 16 02:50:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 10 05:30:07 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Mahesh Dadoda 10/16/2024
 Supervised By : Semsettin Yesilyurt 10/16/2024









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 982
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05

Tgt Ion:168 Resp: 156519
Ion Ratio Lower Upper
168 100
99 67.8 39.1 58.7




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#2
Dichlorodifluoromethane
Concen: 33.800 ug/l
RT: 1.867 min Scan# 24
Delta R.T. 0.001 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01


Tgt Ion: 85 Resp: 49249
Ion Ratio Lower Upper
85 100
87 32.5 16.0 48.0





#3
Chloromethane
Concen: 32.023 ug/l
RT: 2.068 min Scan# 57
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion: 50 Resp: 60729
Ion Ratio Lower Upper
50 100
52 32.6 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024




#4
Vinyl Chloride
Concen: 32.497 ug/l
RT: 2.202 min Scan# 79
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Tgt Ion: 62 Resp: 67816
Ion Ratio Lower Upper
62 100
64 32.5 26.1 39.1





#5
Bromomethane
Concen: 32.242 ug/l
RT: 2.592 min Scan# 144
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion: 94 Resp: 4258
Ion Ratio Lower Upper
94 100
96 94.1 74.1 111.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024




#6
Chloroethane
Concen: 33.401 ug/l
RT: 2.732 min Scan# 166
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01



Tgt Ion: 64 Resp: 46827
Ion Ratio Lower Upper
64 100
66 35.3 26.1 39.1









#7
 Trichlorofluoromethane
 Concen: 34.830 ug/l
 RT: 3.056 min Scan# 219
 Delta R.T. -0.006 min
 Lab File: VY019903.D
 Acq: 15 Oct 2024 15:01

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1015SBS05

Tgt Ion:101 Resp: 10815
 Ion Ratio Lower Upper
 101 100
 103 65.6 49.3 73.9




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
 Supervised By :Semsettin Yesilyurt 10/16/2024



#8
 Diethyl Ether
 Concen: 32.861 ug/l
 RT: 3.458 min Scan# 285
 Delta R.T. 0.000 min
 Lab File: VY019903.D
 Acq: 15 Oct 2024 15:01

Tgt Ion: 74 Resp: 31012
 Ion Ratio Lower Upper
 74 100
 45 103.9 41.3 124.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 35.484 ug/l
RT: 3.818 min Scan# 344
Delta R.T. 0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05

Tgt Ion:101 Resp: 64092
Ion Ratio Lower Upper
101 100
85 48.8 34.5 51.7
151 69.2 61.7 92.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#10
Methyl Iodide
Concen: 30.873 ug/l
RT: 4.007 min Scan# 375
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Tgt Ion:142 Resp: 56398
Ion Ratio Lower Upper
142 100
127 47.7 32.8 49.2
141 13.7 11.4 17.2









#11
Tert butyl alcohol
Concen: 167.379 ug/l
RT: 4.848 min Scan# 514
Delta R.T. -0.012 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05

Tgt Ion: 59 Resp: 30220
Ion Ratio Lower Upper
59 100
57 8.5 3.0 4.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#12
1,1-Dichloroethene
Concen: 31.544 ug/l
RT: 3.793 min Scan# 340
Delta R.T. 0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Tgt Ion: 96 Resp: 52967
Ion Ratio Lower Upper
96 100
61 171.9 114.6 171.8#
98 61.7 50.8 76.2





#13

Acrolein

Concen: 170.996 ug/l

RT: 3.659 min Scan# 317

Delta R.T. 0.006 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Instrument :

MSVOA_Y

ClientSampleId :

VY1015SBS05

Tgt Ion: 56 Resp: 24215

Ion Ratio Lower Upper

56 100





55 71.6 55.7 83.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024

Supervised By :Semsettin Yesilyurt 10/16/2024



#14

Allyl chloride

Concen: 33.674 ug/l

RT: 4.385 min Scan# 437

Delta R.T. 0.000 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01




Tgt Ion: 41 Resp: 102020

Ion Ratio Lower Upper

41 100

39 73.0 58.3 87.5


76 33.7 39.0 58.6#





#15
Acrylonitrile
Concen: 175.163 ug/l
RT: 5.055 min Scan# 547
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05



Tgt Ion: 53 Resp: 7561
Ion Ratio Lower Upper
53 100
52 83.1 64.7 97.1
51 39.0 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#16
Acetone
Concen: 185.047 ug/l
RT: 3.866 min Scan# 352
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Tgt Ion: 43 Resp: 92835
Ion Ratio Lower Upper
43 100
58 28.6 31.7 47.5#









#17
Carbon Disulfide
Concen: 25.687 ug/l
RT: 4.104 min Scan# 391
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05

Tgt Ion: 76 Resp: 11561
Ion Ratio Lower Upper
76 100
78 8.4 6.8 10.2




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#18
Methyl Acetate
Concen: 33.541 ug/l
RT: 4.385 min Scan# 437
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01


Tgt Ion: 43 Resp: 36248
Ion Ratio Lower Upper
43 100
74 21.5 23.3 34.9#





#19
Methyl tert-butyl Ether
Concen: 34.045 ug/l
RT: 5.116 min Scan# 511
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05



Tgt Ion: 73 Resp: 164234
Ion Ratio Lower Upper
73 100
57 22.5 17.3 25.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#20
Methylene Chloride
Concen: 39.987 ug/l
RT: 4.616 min Scan# 475
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Tgt Ion: 84 Resp: 74423
Ion Ratio Lower Upper
84 100
49 131.6 84.4 126.6#
51 39.9 26.0 39.0#
86 59.4 49.5 74.3









#21
trans-1,2-Dichloroethene
Concen: 31.421 ug/l
RT: 5.116 min Scan# 511
Delta R.T. 0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05

Tgt Ion: 96 Resp: 5742
Ion Ratio Lower Upper
96 100
61 140.8 97.5 146.3
98 60.0 50.5 75.7




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#22
Diisopropyl ether
Concen: 35.953 ug/l
RT: 6.018 min Scan# 705
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01


Tgt Ion: 45 Resp: 237258
Ion Ratio Lower Upper
45 100
43 62.8 42.2 63.2
87 24.3 25.0 37.6#
59 9.8 10.5 15.7#





#23
 Vinyl Acetate
 Concen: 171.863 ug/l
 RT: 5.964 min Scan# 696
 Delta R.T. 0.006 min
 Lab File: VY019903.D
 Acq: 15 Oct 2024 15:01




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1015SBS05




Tgt Ion: 43 Resp: 650190
 Ion Ratio Lower Upper
 43 100
 86 9.5 10.6 16.0

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
 Supervised By :Semsettin Yesilyurt 10/16/2024



#24
 1,1-Dichloroethane
 Concen: 34.744 ug/l
 RT: 5.915 min Scan# 688
 Delta R.T. 0.000 min
 Lab File: VY019903.D
 Acq: 15 Oct 2024 15:01




Tgt Ion: 63 Resp: 127452
 Ion Ratio Lower Upper
 63 100
 98 6.8 4.0 12.2
 100 4.8 2.6 8.0





#25
2-Butanone
Concen: 180.159 ug/l
RT: 6.896 min Scan# 849
Delta R.T. 0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05



Tgt Ion: 43 Resp: 120980
Ion Ratio Lower Upper
43 100
72 23.0 27.3 40.9




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#26
2,2-Dichloropropane
Concen: 32.800 ug/l
RT: 6.884 min Scan# 847
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01


Tgt Ion: 77 Resp: 107155
Ion Ratio Lower Upper
77 100
97 21.6 12.0 36.1





#27
cis-1,2-Dichloroethene
Concen: 33.246 ug/l
RT: 6.896 min Scan# 848
Delta R.T. 0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05



Tgt Ion: 96 Resp: 75084
Ion Ratio Lower Upper
96 100
61 159.3 0.0 256.2
98 63.8 0.0 130.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#28
Bromochloromethane
Concen: 32.950 ug/l
RT: 7.250 min Scan# 907
Delta R.T. 0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Tgt Ion: 49 Resp: 53201
Ion Ratio Lower Upper
49 100
129 2.0 0.0 4.6
130 64.4 72.7 109.1#









#29
Tetrahydrofuran
Concen: 172.663 ug/l
RT: 7.268 min Scan# 910
Delta R.T. 0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05

Tgt Ion: 42 Resp: 6620
Ion Ratio Lower Upper
42 100
72 39.5 43.0 64.4
71 38.2 40.8 61.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#30
Chloroform
Concen: 35.354 ug/l
RT: 7.421 min Scan# 935
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01


Tgt Ion: 83 Resp: 132204
Ion Ratio Lower Upper
83 100
85 68.0 52.2 78.4





#31
Cyclohexane
Concen: 31.789 ug/l
RT: 7.701 min Scan# 981
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion: 56 Resp: 10236
Ion Ratio Lower Upper
56 100
69 33.1 28.7 43.1
84 80.8 77.7 116.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#32
1,1,1-Trichloroethane
Concen: 34.601 ug/l
RT: 7.616 min Scan# 967
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01



Tgt Ion: 97 Resp: 113393
Ion Ratio Lower Upper
97 100
99 62.5 50.3 75.5
61 47.5 32.6 48.8





#33

1,2-Dichloroethane-d4

Concen: 72.969 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Instrument :

MSVOA_Y

ClientSampleId :

VY1015SBS05

Tgt Ion: 65 Resp: 140620

Ion Ratio Lower Upper

65 100





67 51.1 0.0 109.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024

Supervised By :Semsettin Yesilyurt 10/16/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01




Tgt Ion:114 Resp: 291370


Ion Ratio Lower Upper

114 100

63 23.0 0.0 35.0


88 16.3 0.0 27.2





#35
Dibromofluoromethane
Concen: 65.318 ug/l
RT: 7.634 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion:113 Resp: 12558
Ion Ratio Lower Upper
113 100
111 104.9 82.2 123.4
192 17.0 15.9 23.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024




#36
1,1-Dichloropropene
Concen: 30.421 ug/l
RT: 7.835 min Scan# 1003
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Tgt Ion: 75 Resp: 84268
Ion Ratio Lower Upper
75 100
110 33.1 19.0 57.0
77 31.7 25.3 37.9





#37
Ethyl Acetate
Concen: 32.033 ug/l
RT: 6.982 min Scan# 80
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05



Tgt Ion: 43 Resp: 46268
Ion Ratio Lower Upper
43 100
61 13.2 12.6 19.0
70 10.3 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#38
Carbon Tetrachloride
Concen: 31.221 ug/l
RT: 7.817 min Scan# 1000
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01


Tgt Ion:117 Resp: 92713
Ion Ratio Lower Upper
117 100
119 99.5 75.8 113.6
121 27.3 24.0 36.0





#39
Methylcyclohexane
Concen: 30.277 ug/l
RT: 9.109 min Scan# 1212
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05



Tgt Ion: 83 Resp: 105679
Ion Ratio Lower Upper
83 100
55 81.4 51.5 77.3
98 48.2 40.5 60.7




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#40
Benzene
Concen: 31.373 ug/l
RT: 8.079 min Scan# 1043
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01


Tgt Ion: 78 Resp: 263002
Ion Ratio Lower Upper
78 100
77 24.2 19.5 29.3





#41
Methacrylonitrile
Concen: 34.564 ug/l
RT: 7.219 min Scan# 902
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion: 41 Resp: 27020
Ion Ratio Lower Upper
41 100
39 54.6 48.0 72.0
67 70.2 75.1 112.7
52 30.1 29.8 44.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024




#42
1,2-Dichloroethane
Concen: 32.504 ug/l
RT: 8.158 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01



Tgt Ion: 62 Resp: 78283
Ion Ratio Lower Upper
62 100
98 9.2 0.0 22.0









#43
Isopropyl Acetate
Concen: 31.459 ug/l
RT: 8.195 min Scan# 1062
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05

Tgt Ion: 43 Resp: 92669
Ion Ratio Lower Upper
43 100
61 17.0 27.9 41.9
87 11.5 13.1 19.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#44
Trichloroethene
Concen: 30.908 ug/l
RT: 8.866 min Scan# 1172
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Tgt Ion:130 Resp: 62756
Ion Ratio Lower Upper
130 100
95 102.3 0.0 188.0





#45

1,2-Dichloropropane

Concen: 32.985 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. 0.000 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Instrument :

MSVOA_Y

ClientSampleId :

VY1015SBS05

Tgt Ion: 63 Resp: 6921

Ion Ratio Lower Upper

63 100





65 31.7 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024

Supervised By :Semsettin Yesilyurt 10/16/2024



#46

Dibromomethane

Concen: 30.276 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. 0.000 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01




Tgt Ion: 93 Resp: 34709


Ion Ratio Lower Upper

93 100

95 85.1 66.2 99.4


174 86.7 80.9 121.3





#47
Bromodichloromethane
Concen: 32.472 ug/l
RT: 9.420 min Scan# 1263
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05



Tgt Ion: 83 Resp: 9881
Ion Ratio Lower Upper
83 100
85 60.1 50.1 75.1
127 8.1 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#48
Methyl methacrylate
Concen: 31.753 ug/l
RT: 9.219 min Scan# 1230
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Tgt Ion: 41 Resp: 41144
Ion Ratio Lower Upper
41 100
69 80.7 84.1 126.1#
39 62.1 43.6 65.4





#49

1,4-Dioxane

Concen: 645.471 ug/l

RT: 9.219 min Scan# 11

Delta R.T. -0.012 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Instrument :

MSVOA_Y

ClientSampleId :





VY1015SBS05

Tgt Ion: 88 Resp: 828
Ion Ratio Lower Upper
88 100
43 39.6 0.0 0.0
58 71.9 49.3 73.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#50

Toluene-d8

Concen: 62.128 ug/l




RT: 10.103 min Scan# 1375


Delta R.T. -0.006 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Tgt Ion: 98 Resp: 441129
Ion Ratio Lower Upper
98 100
100 64.5 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 162.532 ug/l
RT: 9.993 min Scan# 1357
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05

Tgt Ion: 43 Resp: 244204
Ion Ratio Lower Upper
43 100
58 37.0 35.7 53.5




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#52
Toluene
Concen: 31.865 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01


Tgt Ion: 92 Resp: 166746
Ion Ratio Lower Upper
92 100
91 173.1 139.0 208.6





#53
 t-1,3-Dichloropropene
 Concen: 30.662 ug/l
 RT: 10.390 min Scan# 1422
 Delta R.T. -0.006 min
 Lab File: VY019903.D
 Acq: 15 Oct 2024 15:01




Instrument : MSVOA_Y
 ClientSampleId : VY1015SBS05



Tgt Ion: 75 Resp: 83950
 Ion Ratio Lower Upper
 75 100
 77 31.2 24.1 36.1




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
 Supervised By :Semsettin Yesilyurt 10/16/2024



#54
 cis-1,3-Dichloropropene
 Concen: 30.874 ug/l
 RT: 9.853 min Scan# 1334
 Delta R.T. -0.006 min
 Lab File: VY019903.D
 Acq: 15 Oct 2024 15:01


Tgt Ion: 75 Resp: 99743
 Ion Ratio Lower Upper
 75 100
 77 31.0 25.4 38.0
 39 55.3 29.8 44.6#





#55
1,1,2-Trichloroethane
Concen: 32.026 ug/l
RT: 10.573 min Scan# 1452
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion: 97	Resp: 4837
Ion Ratio	Lower Upper
97 100	
83 89.3	67.5 101.3
85 56.4	42.9 64.3
99 60.4	46.6 70.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#56
Ethyl methacrylate
Concen: 31.368 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01



Tgt Ion: 69	Resp: 67984
Ion Ratio	Lower Upper
69 100	
41 71.9	42.3 63.5#
39 43.3	22.9 34.3#





#57

1,3-Dichloropropane

Concen: 32.389 ug/l

RT: 10.713 min Scan# 1475

Delta R.T. -0.006 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Instrument :

MSVOA_Y

ClientSampleId :

VY1015SBS05

Tgt Ion: 76 Resp: 85910

Ion Ratio Lower Upper

76 100





78 31.6 26.2 39.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024

Supervised By :Semsettin Yesilyurt 10/16/2024



#58

2-Chloroethyl Vinyl ether

Concen: 176.572 ug/l

RT: 9.713 min Scan# 1311

Delta R.T. 0.000 min

Lab File: VY019903.D




Acq: 15 Oct 2024 15:01


Tgt Ion: 63 Resp: 178150

Ion Ratio Lower Upper

63 100


106 25.6 23.7 35.5





#59
2-Hexanone
Concen: 167.542 ug/l
RT: 10.762 min Scan# 1482
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01



Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05



Tgt Ion: 43 Resp: 179798
Ion Ratio Lower Upper
43 100
58 52.0 31.4 94.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#60
Dibromochloromethane
Concen: 31.464 ug/l
RT: 10.908 min Scan# 1507
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01


Tgt Ion:129 Resp: 60998
Ion Ratio Lower Upper
129 100
127 77.7 39.0 117.0





#61
1,2-Dibromoethane
Concen: 31.941 ug/l
RT: 11.011 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion:107 Resp: 43590
Ion Ratio Lower Upper
107 100
109 90.4 75.9 113.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024




#62
4-Bromofluorobenzene
Concen: 64.716 ug/l
RT: 12.401 min Scan# 1752
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Tgt Ion: 95 Resp: 166034
Ion Ratio Lower Upper
95 100
174 73.2 0.0 175.6
176 70.7 0.0 171.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion:117 Resp: 270640
Ion Ratio Lower Upper
117 100
82 61.3 42.4 63.6
119 33.2 25.9 38.9

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024




#64
Tetrachloroethene
Concen: 27.683 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Tgt Ion:164 Resp: 53292
Ion Ratio Lower Upper
164 100
166 126.6 102.6 154.0
129 98.3 72.0 108.0
131 95.1 70.0 105.0





#65
Chlorobenzene
Concen: 29.106 ug/l
RT: 11.438 min Scan# 11594
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion:112 Resp: 180178
Ion Ratio Lower Upper
112 100
114 31.6 26.4 39.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 29.436 ug/l
RT: 11.511 min Scan# 1606
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Tgt Ion:131 Resp: 61650
Ion Ratio Lower Upper
131 100
133 96.0 47.8 143.3
119 66.2 34.1 102.2





#67
Ethyl Benzene
Concen: 29.188 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion: 91 Resp: 328219
Ion Ratio Lower Upper
91 100
106 30.2 24.5 36.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024




#68
m/p-Xylenes
Concen: 57.885 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Tgt Ion:106 Resp: 240397
Ion Ratio Lower Upper
106 100
91 206.2 156.0 234.0





#69
o-Xylene
Concen: 29.276 ug/l
RT: 11.950 min Scan# 1678
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion:106 Resp: 116819
Ion Ratio Lower Upper
106 100
91 218.3 103.6 310.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#70
Styrene
Concen: 29.067 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01



Tgt Ion:104 Resp: 195685
Ion Ratio Lower Upper
104 100
78 52.9 37.0 55.4
103 54.6 44.2 66.2





#71

Bromoform

Concen: 28.709 ug/l

RT: 12.127 min Scan# 1707

Delta R.T. -0.006 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Instrument :

MSVOA_Y

ClientSampleId :





VY1015SBS05

Tgt Ion:	173	Resp:	3218
Ion Ratio	Lower	Upper	
173	100		
175	47.1	25.1	75.3
254	0.0	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.346 min Scan# 1907


Delta R.T. 0.000 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Tgt Ion:	152	Resp:	138870
Ion Ratio	Lower	Upper	
152	100		
115	62.9	28.2	84.7
150	167.6	0.0	345.6





#73

Isopropylbenzene

Concen: 27.264 ug/l

RT: 12.249 min Scan# 11

Delta R.T. -0.006 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Instrument :

MSVOA_Y

ClientSampleId :

VY1015SBS05

Tgt Ion:105 Resp: 31846

Ion Ratio Lower Upper

105 100





120 25.5 13.4 40.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024

Supervised By :Semsettin Yesilyurt 10/16/2024



#74

N-ethyl acetate

Concen: 26.047 ug/l

RT: 12.066 min Scan# 1697

Delta R.T. -0.006 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Tgt Ion: 43 Resp: 80584




Ion Ratio Lower Upper


43 100

70 40.7 41.4 62.2#

55 23.7 23.3 34.9

61 23.2 22.6 33.8









#75
1,1,2,2-Tetrachloroethane
Concen: 28.107 ug/l
RT: 12.505 min Scan# 1769
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05

Tgt Ion: 83 Resp: 5833
Ion Ratio Lower Upper
83 100
131 9.9 5.0 14.9
85 64.8 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#76
1,2,3-Trichloropropane
Concen: 14.024 ug/l m
RT: 12.554 min Scan# 1777
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Tgt Ion: 75 Resp: 38151
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 26.659 ug/l

RT: 12.530 min Scan# 1773

Delta R.T. -0.006 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Instrument :

MSVOA_Y

ClientSampleId :

VY1015SBS05





Tgt Ion:	156	Resp:	66080
Ion Ratio	Lower	Upper	
156	100		
77	208.1	87.1	261.3
158	96.0	48.6	145.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024

Supervised By :Semsettin Yesilyurt 10/16/2024



#78

n-propylbenzene

Concen: 28.111 ug/l




RT: 12.590 min Scan# 1783


Delta R.T. -0.006 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Tgt Ion:	91	Resp:	396766
Ion Ratio	Lower	Upper	
91	100		
120	20.9	11.5	34.4









#79
2-Chlorotoluene
Concen: 27.732 ug/l
RT: 12.676 min Scan# 1797
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01

Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05

Tgt Ion: 91 Resp: 22439
Ion Ratio Lower Upper
91 100
126 31.6 17.8 53.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#80
1,3,5-Trimethylbenzene
Concen: 27.489 ug/l
RT: 12.731 min Scan# 1806
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01


Tgt Ion:105 Resp: 259309
Ion Ratio Lower Upper
105 100
120 47.2 25.1 75.2





#81
trans-1,4-Dichloro-2-butene
Concen: 24.993 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05



Tgt Ion: 75 Resp: 1684
Ion Ratio Lower Upper
75 100
53 109.5 66.8 100.2
89 45.4 35.8 53.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#82
4-Chlorotoluene
Concen: 27.076 ug/l
RT: 12.773 min Scan# 1813
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01


Tgt Ion: 91 Resp: 224020
Ion Ratio Lower Upper
91 100
126 33.5 17.6 52.9





#83
tert-Butylbenzene
Concen: 27.750 ug/l
RT: 12.993 min Scan# 1849
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion:119 Resp: 234509
Ion Ratio Lower Upper
119 100
91 70.2 31.3 93.8
134 27.3 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024




#84
1,2,4-Trimethylbenzene
Concen: 27.315 ug/l
RT: 13.042 min Scan# 1857
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01



Tgt Ion:105 Resp: 255465
Ion Ratio Lower Upper
105 100
120 44.2 23.2 69.6









#85
 sec-Butylbenzene
 Concen: 28.595 ug/l
 RT: 13.170 min Scan# 1878
 Delta R.T. -0.006 min
 Lab File: VY019903.D
 Acq: 15 Oct 2024 15:01

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1015SBS05

Tgt Ion:105 Resp: 361040
 Ion Ratio Lower Upper
 105 100
 134 18.6 10.1 30.1




Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
 Supervised By :Semsettin Yesilyurt 10/16/2024



#86
 p-Isopropyltoluene
 Concen: 27.866 ug/l
 RT: 13.285 min Scan# 1897
 Delta R.T. -0.006 min
 Lab File: VY019903.D
 Acq: 15 Oct 2024 15:01

Tgt Ion:119 Resp: 285701
 Ion Ratio Lower Upper
 119 100
 134 25.6 13.1 39.3
 91 24.8 10.7 32.1





#87

1,3-Dichlorobenzene

Concen: 27.233 ug/l

RT: 13.285 min Scan# 1897

Delta R.T. -0.006 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01

Instrument :

MSVOA_Y

ClientSampleId :

VY1015SBS05

Tgt Ion:146 Resp: 136318

Ion Ratio Lower Upper

146 100

111 42.0 19.0 57.0





148 63.2 31.7 95.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024

Supervised By :Semsettin Yesilyurt 10/16/2024



#88

1,4-Dichlorobenzene

Concen: 26.964 ug/l

RT: 13.365 min Scan# 1910

Delta R.T. -0.006 min

Lab File: VY019903.D

Acq: 15 Oct 2024 15:01




Tgt Ion:146 Resp: 132600


Ion Ratio Lower Upper

146 100

111 42.9 18.4 55.4


148 62.3 32.4 97.2





#89
n-Butylbenzene
Concen: 28.391 ug/l
RT: 13.615 min Scan# 1951
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05



Tgt Ion: 91 Resp: 284233
Ion Ratio Lower Upper
91 100
92 53.4 26.6 79.7
134 23.3 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#90
Hexachloroethane
Concen: 27.176 ug/l
RT: 13.877 min Scan# 1994
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01


Tgt Ion:117 Resp: 54255
Ion Ratio Lower Upper
117 100
201 64.0 37.4 112.2





#91
1,2-Dichlorobenzene
Concen: 27.200 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion:146 Resp: 11963
Ion Ratio Lower Upper
146 100
111 43.3 19.9 59.7
148 63.3 31.8 95.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 27.261 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Tgt Ion: 75 Resp: 9030
Ion Ratio Lower Upper
75 100
155 69.6 48.6 145.8
157 92.3 62.7 188.2





#93
1,2,4-Trichlorobenzene
Concen: 26.559 ug/l
RT: 14.919 min Scan# 2165
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion:180 Resp: 65502
Ion Ratio Lower Upper
180 100
182 94.8 47.9 143.8
145 34.1 14.2 42.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024




#94
Hexachlorobutadiene
Concen: 27.324 ug/l
RT: 15.023 min Scan# 2182
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Tgt Ion:225 Resp: 37465
Ion Ratio Lower Upper
225 100
223 64.2 30.4 91.2
227 61.4 31.9 95.9





#95
Naphthalene
Concen: 26.710 ug/l
RT: 15.139 min Scan# 21
Delta R.T. -0.006 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01




Instrument :
MSVOA_Y
ClientSampleId :
VY1015SBS05




Tgt Ion:128 Resp: 124244
Ion Ratio Lower Upper
128 100
127 12.6 10.6 16.0
129 11.5 8.6 13.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 10/16/2024
Supervised By :Semsettin Yesilyurt 10/16/2024



#96
1,2,3-Trichlorobenzene
Concen: 26.450 ug/l
RT: 15.328 min Scan# 2232
Delta R.T. 0.000 min
Lab File: VY019903.D
Acq: 15 Oct 2024 15:01



Tgt Ion:180 Resp: 55142
Ion Ratio Lower Upper
180 100
182 95.6 47.6 142.9
145 33.2 15.4 46.1

