

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101625\
 Data File : VY023482.D
 Acq On : 16 Oct 2025 09:46
 Operator : SY/MD
 Sample : VY1016SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1016SBL01

Quant Time: Oct 17 01:51:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	529968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	1064919	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1102445	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	465964	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	279369	63.039	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.080%
35) Dibromofluoromethane	7.640	113	359646	54.963	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.920%
50) Toluene-d8	10.109	98	1301045	53.391	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.780%
62) 4-Bromofluorobenzene	12.408	95	430344	53.491	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101625\
Data File : VY023482.D
Acq On : 16 Oct 2025 09:46
Operator : SY/MD
Sample : VY1016SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1016SBL01

Quant Time: Oct 17 01:51:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

