

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101625\
 Data File : VY023493.D
 Acq On : 16 Oct 2025 16:27
 Operator : SY/MD
 Sample : Q3367-01
 Misc : 4.32g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BU-03-101625

Quant Time: Oct 17 01:59:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	632316	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1258574	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1305396	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	604308	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	311218	58.859	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.720%
35) Dibromofluoromethane	7.640	113	416785	53.895	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.780%
50) Toluene-d8	10.109	98	1531486	53.177	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.360%
62) 4-Bromofluorobenzene	12.408	95	534883	56.255	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	112.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101625\
Data File : VY023493.D
Acq On : 16 Oct 2025 16:27
Operator : SY/MD
Sample : Q3367-01
Misc : 4.32g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BU-03-101625

Quant Time: Oct 17 01:59:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

