

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102025\
 Data File : VY023499.D
 Acq On : 20 Oct 2025 11:14
 Operator : SY/MD
 Sample : Q3390-03
 Misc : 1.00g/5.0mL/MSVOA_Y/SOIL/C
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 4174

Quant Time: Oct 21 03:56:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	780680	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1295443	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1069873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	398027	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	352999	54.073	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.140%
35) Dibromofluoromethane	7.634	113	408822	51.360	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.720%
50) Toluene-d8	10.109	98	1428367	48.185	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.380%
62) 4-Bromofluorobenzene	12.408	95	384524	39.290	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102025\
Data File : VY023499.D
Acq On : 20 Oct 2025 11:14
Operator : SY/MD
Sample : Q3390-03
Misc : 1.00g/5.0mL/MSVOA_Y/SOIL/C
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
4174

Quant Time: Oct 21 03:56:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

