

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102025\
 Data File : VY023509.D
 Acq On : 20 Oct 2025 15:23
 Operator : SY/MD
 Sample : Q3392-03
 Misc : 3.85g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-20-VOC

Quant Time: Oct 21 03:59:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	684849	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1177232	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	995247	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	411384	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	329934	57.612	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.220%
35) Dibromofluoromethane	7.640	113	387381	53.553	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.100%
50) Toluene-d8	10.109	98	1347428	50.019	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.040%
62) 4-Bromofluorobenzene	12.408	95	365399	41.085	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.180%
Target Compounds						
16) Acetone	3.873	43	10099	7.407	ug/l	Qvalue # 75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102025\
Data File : VY023509.D
Acq On : 20 Oct 2025 15:23
Operator : SY/MD
Sample : Q3392-03
Misc : 3.85g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-20-VOC

Quant Time: Oct 21 03:59:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

