

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102124\
 Data File : VY019958.D
 Acq On : 21 Oct 2024 12:50
 Operator : SY/MD
 Sample : P4443-02
 Misc : 4.52g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OG-315-HR-502-VOC-29

Quant Time: Oct 21 23:30:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	216896	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	446253	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	403428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	133778	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	149122	55.838	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 111.680%	
35) Dibromofluoromethane	7.634	113	146044	49.594	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.180%	
50) Toluene-d8	10.103	98	554220	50.964	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.920%	
62) 4-Bromofluorobenzene	12.407	95	162642	41.392	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 82.780%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102124\
Data File : VY019958.D
Acq On : 21 Oct 2024 12:50
Operator : SY/MD
Sample : P4443-02
Misc : 4.52g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OG-315-HR-502-VOC-29

Quant Time: Oct 21 23:30:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

