

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102124\
 Data File : VY019954.D
 Acq On : 21 Oct 2024 10:38
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 21 23:28:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	65	0.00
2 T	Dichlorodifluoromethane	0.465	0.428	8.0	61	0.00
3 P	Chloromethane	0.606	0.617	-1.8	65	0.00
4 C	Vinyl Chloride	0.667	0.722	-8.2#	68	0.00
5 T	Bromomethane	0.422	0.450	-6.6	68	0.00
6 T	Chloroethane	0.448	0.497	-10.9	71	0.00
7 T	Trichlorofluoromethane	0.992	1.057	-6.6	69	0.00
8 T	Diethyl Ether	0.301	0.296	1.7	62	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.577	0.636	-10.2	71	0.00
10 T	Methyl Iodide	0.584	0.573	1.9	59	0.00
11 T	Tert butyl alcohol	0.058	0.039	32.8#	51	0.01
12 CM	1,1-Dichloroethene	0.536	0.547	-2.1#	65	0.00
13 T	Acrolein	0.045	0.029	35.6#	47#	0.00
14 T	Allyl chloride	0.968	1.046	-8.1	67	0.00
15 T	Acrylonitrile	0.138	0.146	-5.8	66	0.00
16 T	Acetone	0.160	0.172	-7.5	66	0.00
17 T	Carbon Disulfide	1.438	1.271	11.6	53	0.00
18 T	Methyl Acetate	0.345	0.359	-4.1	67	0.00
19 T	Methyl tert-butyl Ether	1.541	1.595	-3.5	66	0.00
20 T	Methylene Chloride	0.647	0.627	3.1	66	0.00
21 T	trans-1,2-Dichloroethene	0.584	0.617	-5.7	66	0.00
22 T	Diisopropyl ether	2.108	2.413	-14.5	73	0.00
23 T	Vinyl Acetate	1.209	1.296	-7.2	66	0.00
24 P	1,1-Dichloroethane	1.172	1.319	-12.5	71	0.00
25 T	2-Butanone	0.215	0.222	-3.3	65	0.00
26 T	2,2-Dichloropropane	1.044	1.145	-9.7	71	0.00
27 T	cis-1,2-Dichloroethene	0.721	0.771	-6.9	68	0.00
28 T	Bromochloromethane	0.516	0.606	-17.4	74	0.00
29 T	Tetrahydrofuran	0.122	0.122	0.0	62	0.00
30 C	Chloroform	1.195	1.363	-14.1#	72	0.00
31 T	Cyclohexane	1.029	1.032	-0.3	67	0.00
32 T	1,1,1-Trichloroethane	1.047	1.171	-11.8	71	0.00
33 S	1,2-Dichloroethane-d4	0.616	0.616	0.0	69	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	65	0.00
35 S	Dibromofluoromethane	0.330	0.352	-6.7	71	0.00
36 T	1,1-Dichloropropene	0.475	0.528	-11.2	69	0.00
37 T	Ethyl Acetate	0.248	0.251	-1.2	63	0.00
38 T	Carbon Tetrachloride	0.510	0.572	-12.2	71	0.00
39 T	Methylcyclohexane	0.599	0.636	-6.2	66	0.00
40 TM	Benzene	1.439	1.584	-10.1	69	0.00
41 T	Methacrylonitrile	0.134	0.151	-12.7	63	0.00
42 TM	1,2-Dichloroethane	0.413	0.451	-9.2	68	0.00
43 T	Isopropyl Acetate	0.505	0.520	-3.0	65	0.00
44 TM	Trichloroethene	0.348	0.374	-7.5	66	0.00
45 C	1,2-Dichloropropane	0.360	0.412	-14.4#	73	0.00
46 T	Dibromomethane	0.197	0.208	-5.6	66	0.00
47 T	Bromodichloromethane	0.522	0.594	-13.8	71	0.00
48 T	Methyl methacrylate	0.222	0.237	-6.8	65	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102124\
 Data File : VY019954.D
 Acq On : 21 Oct 2024 10:38
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 21 23:28:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	62	0.00
50 S	Toluene-d8	1.218	1.289	-5.8	71	0.00
51 T	4-Methyl-2-Pentanone	0.258	0.275	-6.6	67	0.00
52 CM	Toluene	0.898	1.021	-13.7#	71	0.00
53 T	t-1,3-Dichloropropene	0.470	0.508	-8.1	66	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.610	-10.1	69	0.00
55 T	1,1,2-Trichloroethane	0.259	0.279	-7.7	68	0.00
56 T	Ethyl methacrylate	0.372	0.394	-5.9	64	0.00
57 T	1,3-Dichloropropane	0.455	0.494	-8.6	68	0.00
58 T	2-Chloroethyl Vinyl ether	0.173	0.187	-8.1	71	0.00
59 T	2-Hexanone	0.184	0.199	-8.2	66	0.00
60 T	Dibromochloromethane	0.333	0.356	-6.9	68	0.00
61 T	1,2-Dibromoethane	0.234	0.242	-3.4	64	0.00
62 S	4-Bromofluorobenzene	0.440	0.462	-5.0	71	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	65	0.00
64 T	Tetrachloroethene	0.356	0.367	-3.1	64	0.00
65 PM	Chlorobenzene	1.144	1.258	-10.0	69	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.434	-12.1	71	0.00
67 C	Ethyl Benzene	2.077	2.378	-14.5#	71	0.00
68 T	m/p-Xylenes	0.767	0.876	-14.2	71	0.00
69 T	o-Xylene	0.737	0.840	-14.0	71	0.00
70 T	Styrene	1.244	1.419	-14.1	70	0.00
71 P	Bromoform	0.207	0.222	-7.2	65	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	65	0.00
73 T	Isopropylbenzene	4.206	4.833	-14.9	73	0.00
74 T	N-amyl acetate	1.114	1.201	-7.8	66	0.00
75 P	1,1,2,2-Tetrachloroethane	0.747	0.825	-10.4	70	0.00
76 T	1,2,3-Trichloropropane	0.979	0.553	43.5#	36#	0.00
77 T	Bromobenzene	0.892	0.953	-6.8	68	0.00
78 T	n-propylbenzene	5.082	5.962	-17.3	74	0.00
79 T	2-Chlorotoluene	2.913	3.277	-12.5	71	0.00
80 T	1,3,5-Trimethylbenzene	3.396	3.848	-13.3	72	0.00
81 T	trans-1,4-Dichloro-2-butene	0.243	0.259	-6.6	67	0.00
82 T	4-Chlorotoluene	2.979	3.374	-13.3	72	0.00
83 T	tert-Butylbenzene	3.043	3.576	-17.5	75	0.00
84 T	1,2,4-Trimethylbenzene	3.367	3.840	-14.0	72	0.00
85 T	sec-Butylbenzene	4.546	5.428	-19.4	74	0.00
86 T	p-Isopropyltoluene	3.691	4.301	-16.5	73	0.00
87 T	1,3-Dichlorobenzene	1.802	1.956	-8.5	70	0.00
88 T	1,4-Dichlorobenzene	1.771	1.909	-7.8	69	0.00
89 T	n-Butylbenzene	3.605	4.338	-20.3	73	0.00
90 T	Hexachloroethane	0.719	0.861	-19.7	75	0.00
91 T	1,2-Dichlorobenzene	1.584	1.718	-8.5	69	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.119	0.120	-0.8	64	0.00
93 T	1,2,4-Trichlorobenzene	0.888	0.947	-6.6	63	0.00
94 T	Hexachlorobutadiene	0.494	0.552	-11.7	67	0.00
95 T	Naphthalene	1.675	1.720	-2.7	60	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102124\
Data File : VY019954.D
Acq On : 21 Oct 2024 10:38
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 21 23:28:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.751	0.787	-4.8	62	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6