

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102125\
 Data File : VY023525.D
 Acq On : 21 Oct 2025 11:03
 Operator : SY/MD
 Sample : Q3383-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-03-101725

Quant Time: Oct 22 14:12:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	784527	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1309945	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1102151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	445690	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	355143	54.135	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.260%
35) Dibromofluoromethane	7.640	113	420121	52.195	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.400%
50) Toluene-d8	10.109	98	1458229	48.648	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.300%
62) 4-Bromofluorobenzene	12.408	95	420899	42.531	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102125\
Data File : VY023525.D
Acq On : 21 Oct 2025 11:03
Operator : SY/MD
Sample : Q3383-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-101725

Quant Time: Oct 22 14:12:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

