

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102125\
 Data File : VY023534.D
 Acq On : 21 Oct 2025 14:59
 Operator : SY/MD
 Sample : Q3399-03
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-ROLLOFF

Quant Time: Oct 22 12:28:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 07 14:25:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	803329	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1303553	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1099444	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	457774	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	336669	50.118	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.240%
35) Dibromofluoromethane	7.640	113	409395	51.112	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.220%
50) Toluene-d8	10.109	98	1472481	49.364	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.720%
62) 4-Bromofluorobenzene	12.408	95	421770	42.828	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102125\
Data File : VY023534.D
Acq On : 21 Oct 2025 14:59
Operator : SY/MD
Sample : Q3399-03
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-ROLLOFF

Quant Time: Oct 22 12:28:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 07 14:25:11 2025
Response via : Initial Calibration

