

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102224\
 Data File : VY019974.D
 Acq On : 22 Oct 2024 12:26
 Operator : SY/MD
 Sample : P4467-03
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1-VOC

Quant Time: Oct 23 01:27:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	217854	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	458881	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	427187	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	148963	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	160442	59.812	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 119.620%	
35) Dibromofluoromethane	7.634	113	154596	51.054	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.100%	
50) Toluene-d8	10.103	98	568032	50.797	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.600%	
62) 4-Bromofluorobenzene	12.402	95	177789	44.001	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 88.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102224\
Data File : VY019974.D
Acq On : 22 Oct 2024 12:26
Operator : SY/MD
Sample : P4467-03
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1-VOC

Quant Time: Oct 23 01:27:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

