

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
 Data File : VY023548.D
 Acq On : 22 Oct 2025 12:03
 Operator : SY/MD
 Sample : Q3416-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-08-10212025

Quant Time: Oct 23 13:13:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	956693	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	1547678	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1323539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	562005	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	384980	48.122	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.240%
35) Dibromofluoromethane	7.640	113	468225	49.236	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.480%
50) Toluene-d8	10.109	98	1759871	49.692	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.380%
62) 4-Bromofluorobenzene	12.407	95	523647	44.786	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.580%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.622	84	12259	1.190	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102225\
Data File : VY023548.D
Acq On : 22 Oct 2025 12:03
Operator : SY/MD
Sample : Q3416-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-08-10212025

Quant Time: Oct 23 13:13:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

