

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102324\
 Data File : VY019990.D
 Acq On : 23 Oct 2024 13:35
 Operator : SY/MD
 Sample : P4487-05
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-F27

Quant Time: Oct 23 23:07:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	198616	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	403061	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	366132	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	130434	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	145586	59.531	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.060%
35) Dibromofluoromethane	7.640	113	136023	51.141	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.280%
50) Toluene-d8	10.103	98	502223	51.132	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.260%
62) 4-Bromofluorobenzene	12.401	95	153800	43.336	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	86.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102324\
Data File : VY019990.D
Acq On : 23 Oct 2024 13:35
Operator : SY/MD
Sample : P4487-05
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-F27

Quant Time: Oct 23 23:07:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

