

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102324\
 Data File : VY019985.D
 Acq On : 23 Oct 2024 10:58
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 23 23:03:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	66	0.00
2 T	Dichlorodifluoromethane	0.465	0.411	11.6	59	0.00
3 P	Chloromethane	0.606	0.604	0.3	65	0.00
4 C	Vinyl Chloride	0.667	0.694	-4.0#	66	0.00
5 T	Bromomethane	0.422	0.448	-6.2	68	0.00
6 T	Chloroethane	0.448	0.492	-9.8	71	0.00
7 T	Trichlorofluoromethane	0.992	1.072	-8.1	71	0.00
8 T	Diethyl Ether	0.301	0.314	-4.3	66	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.577	0.622	-7.8	71	0.00
10 T	Methyl Iodide	0.584	0.554	5.1	58	0.00
11 T	Tert butyl alcohol	0.058	0.046	20.7	61	0.00
12 CM	1,1-Dichloroethene	0.536	0.538	-0.4#	65	0.00
13 T	Acrolein	0.045	0.028	37.8#	46#	0.00
14 T	Allyl chloride	0.968	1.056	-9.1	69	0.00
15 T	Acrylonitrile	0.138	0.150	-8.7	69	0.00
16 T	Acetone	0.160	0.176	-10.0	68	0.00
17 T	Carbon Disulfide	1.438	1.233	14.3	52	0.00
18 T	Methyl Acetate	0.345	0.377	-9.3	71	0.00
19 T	Methyl tert-butyl Ether	1.541	1.661	-7.8	69	0.00
20 T	Methylene Chloride	0.647	0.640	1.1	68	0.00
21 T	trans-1,2-Dichloroethene	0.584	0.603	-3.3	66	0.00
22 T	Diisopropyl ether	2.108	2.477	-17.5	75	0.00
23 T	Vinyl Acetate	1.209	1.338	-10.7	69	0.00
24 P	1,1-Dichloroethane	1.172	1.359	-16.0	74	0.00
25 T	2-Butanone	0.215	0.227	-5.6	67	0.00
26 T	2,2-Dichloropropane	1.044	1.162	-11.3	73	0.00
27 T	cis-1,2-Dichloroethene	0.721	0.807	-11.9	72	0.00
28 T	Bromochloromethane	0.516	0.604	-17.1	75	0.00
29 T	Tetrahydrofuran	0.122	0.128	-4.9	66	0.00
30 C	Chloroform	1.195	1.409	-17.9#	76	0.00
31 T	Cyclohexane	1.029	1.019	1.0	66	0.00
32 T	1,1,1-Trichloroethane	1.047	1.201	-14.7	74	0.00
33 S	1,2-Dichloroethane-d4	0.616	0.728	-18.2	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	68	0.00
35 S	Dibromofluoromethane	0.330	0.395	-19.7	84	0.00
36 T	1,1-Dichloropropene	0.475	0.519	-9.3	71	0.00
37 T	Ethyl Acetate	0.248	0.254	-2.4	67	0.00
38 T	Carbon Tetrachloride	0.510	0.557	-9.2	72	0.00
39 T	Methylcyclohexane	0.599	0.618	-3.2	67	0.00
40 TM	Benzene	1.439	1.569	-9.0	72	0.00
41 T	Methacrylonitrile	0.134	0.134	0.0	59	0.00
42 TM	1,2-Dichloroethane	0.413	0.455	-10.2	72	0.00
43 T	Isopropyl Acetate	0.505	0.524	-3.8	68	0.00
44 TM	Trichloroethene	0.348	0.364	-4.6	68	0.00
45 C	1,2-Dichloropropane	0.360	0.404	-12.2#	75	0.00
46 T	Dibromomethane	0.197	0.204	-3.6	68	0.00
47 T	Bromodichloromethane	0.522	0.598	-14.6	75	0.00
48 T	Methyl methacrylate	0.222	0.240	-8.1	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102324\
 Data File : VY019985.D
 Acq On : 23 Oct 2024 10:58
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 23 23:03:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	63	0.00
50 S	Toluene-d8	1.218	1.417	-16.3	81	0.00
51 T	4-Methyl-2-Pentanone	0.258	0.275	-6.6	69	0.00
52 CM	Toluene	0.898	0.993	-10.6#	72	0.00
53 T	t-1,3-Dichloropropene	0.470	0.510	-8.5	70	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.608	-9.7	72	0.00
55 T	1,1,2-Trichloroethane	0.259	0.272	-5.0	69	0.00
56 T	Ethyl methacrylate	0.372	0.398	-7.0	67	0.00
57 T	1,3-Dichloropropane	0.455	0.493	-8.4	71	0.00
58 T	2-Chloroethyl Vinyl ether	0.173	0.185	-6.9	74	0.00
59 T	2-Hexanone	0.184	0.198	-7.6	68	0.00
60 T	Dibromochloromethane	0.333	0.364	-9.3	72	0.00
61 T	1,2-Dibromoethane	0.234	0.242	-3.4	67	0.00
62 S	4-Bromofluorobenzene	0.440	0.515	-17.0	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	67	0.00
64 T	Tetrachloroethene	0.356	0.354	0.6	65	0.00
65 PM	Chlorobenzene	1.144	1.251	-9.4	72	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.433	-11.9	73	0.00
67 C	Ethyl Benzene	2.077	2.355	-13.4#	73	0.00
68 T	m/p-Xylenes	0.767	0.864	-12.6	73	0.00
69 T	o-Xylene	0.737	0.825	-11.9	72	0.00
70 T	Styrene	1.244	1.413	-13.6	73	0.00
71 P	Bromoform	0.207	0.221	-6.8	67	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	67	0.00
73 T	Isopropylbenzene	4.206	4.708	-11.9	74	0.00
74 T	N-amyl acetate	1.114	1.177	-5.7	68	0.00
75 P	1,1,2,2-Tetrachloroethane	0.747	0.789	-5.6	70	0.00
76 T	1,2,3-Trichloropropane	0.979	0.570	41.8#	39#	0.00
77 T	Bromobenzene	0.892	0.920	-3.1	68	0.00
78 T	n-propylbenzene	5.082	5.789	-13.9	75	0.00
79 T	2-Chlorotoluene	2.913	3.229	-10.8	73	0.00
80 T	1,3,5-Trimethylbenzene	3.396	3.783	-11.4	74	0.00
81 T	trans-1,4-Dichloro-2-butene	0.243	0.250	-2.9	67	0.00
82 T	4-Chlorotoluene	2.979	3.331	-11.8	74	0.00
83 T	tert-Butylbenzene	3.043	3.463	-13.8	75	0.00
84 T	1,2,4-Trimethylbenzene	3.367	3.771	-12.0	73	0.00
85 T	sec-Butylbenzene	4.546	5.230	-15.0	75	0.00
86 T	p-Isopropyltoluene	3.691	4.145	-12.3	73	0.00
87 T	1,3-Dichlorobenzene	1.802	1.923	-6.7	71	0.00
88 T	1,4-Dichlorobenzene	1.771	1.886	-6.5	71	0.00
89 T	n-Butylbenzene	3.605	4.208	-16.7	74	0.00
90 T	Hexachloroethane	0.719	0.820	-14.0	75	0.00
91 T	1,2-Dichlorobenzene	1.584	1.705	-7.6	71	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.119	0.119	0.0	66	0.00
93 T	1,2,4-Trichlorobenzene	0.888	0.935	-5.3	65	0.00
94 T	Hexachlorobutadiene	0.494	0.515	-4.3	66	0.00
95 T	Naphthalene	1.675	1.700	-1.5	62	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102324\
Data File : VY019985.D
Acq On : 23 Oct 2024 10:58
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 23 23:03:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.751	0.768	-2.3	63	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6