

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102422\
 Data File : VY011066.D
 Acq On : 24 Oct 2022 14:05
 Operator : KP/MD
 Sample : N5217-23
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-E3-7.5-8.0

Quant Time: Oct 25 05:29:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	7316	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	11317	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	9495	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	2919	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	5207	77.265	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	154.540%
35) Dibromofluoromethane	7.709	113	2668	41.946	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.900%
50) Toluene-d8	10.178	98	14413	60.995	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	121.980%
62) 4-Bromofluorobenzene	12.483	95	3917	46.586	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102422\
Data File : VY011066.D
Acq On : 24 Oct 2022 14:05
Operator : KP/MD
Sample : N5217-23
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-E3-7.5-8.0

Quant Time: Oct 25 05:29:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

