

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102422\
 Data File : VY011067.D
 Acq On : 24 Oct 2022 14:28
 Operator : KP/MD
 Sample : N5254-01
 Misc : 7.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1-102122

Quant Time: Oct 25 05:29:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	296550	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	500280	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	409038	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	150736	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	140438	51.411	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.820%	
35) Dibromofluoromethane	7.710	113	151148	54.389	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.780%	
50) Toluene-d8	10.179	98	515532	49.353	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 98.700%	
62) 4-Bromofluorobenzene	12.477	95	145366	38.584	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 77.160%	
Target Compounds						
16) Acetone	3.948	43	2150	2.710	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102422\
Data File : VY011067.D
Acq On : 24 Oct 2022 14:28
Operator : KP/MD
Sample : N5254-01
Misc : 7.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1-102122

Quant Time: Oct 25 05:29:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

