

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102422\
 Data File : VY011072.D
 Acq On : 24 Oct 2022 16:21
 Operator : KP/MD
 Sample : N5253-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 164718

Quant Time: Oct 25 05:47:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	45033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	72247	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	65181	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	25489	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	29470	71.043	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	142.080%
35) Dibromofluoromethane	7.710	113	28279	71.130	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	142.260%
50) Toluene-d8	10.179	98	85103	56.415	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.820%
62) 4-Bromofluorobenzene	12.477	95	26218	49.003	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102422\
Data File : VY011072.D
Acq On : 24 Oct 2022 16:21
Operator : KP/MD
Sample : N5253-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
164718

Quant Time: Oct 25 05:47:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

