

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102422\
 Data File : VY011076.D
 Acq On : 24 Oct 2022 17:52
 Operator : KP/MD
 Sample : N5238-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-20430

Quant Time: Oct 25 05:47:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	275393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	470279	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	398274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	172418	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	156250	61.594	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.180%
35) Dibromofluoromethane	7.716	113	167610	64.565	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	129.140%
50) Toluene-d8	10.179	98	568151	57.860	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.720%
62) 4-Bromofluorobenzene	12.477	95	183507	52.938	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.880%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102422\
Data File : VY011076.D
Acq On : 24 Oct 2022 17:52
Operator : KP/MD
Sample : N5238-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CHRT-20430

Quant Time: Oct 25 05:47:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY011076.D
Acq: 24 Oct 2022 17:52




Instrument :
MSVOA_Y
ClientSampleId :
CHRT-20430


Tgt Ion:168 Resp: 275393
Ion Ratio Lower Upper
168 100
99 54.4 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 61.594 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY011076.D
Acq: 24 Oct 2022 17:52

Tgt Ion: 65 Resp: 156250
Ion Ratio Lower Upper
65 100
67 54.8 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY011076.D

Acq: 24 Oct 2022 17:52

Instrument :

MSVOA_Y

ClientSampleId :

CHRT-20430





Tgt Ion:114 Resp: 470279

Ion Ratio Lower Upper

114 100

63 19.6 0.0 43.2

88 15.7 0.0 31.8



#35

Dibromofluoromethane

Concen: 64.565 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.006 min

Lab File: VY011076.D

Acq: 24 Oct 2022 17:52




Tgt Ion:113 Resp: 167610


Ion Ratio Lower Upper

113 100

111 102.9 82.6 124.0

192 19.4 17.6 26.4









#50
Toluene-d8
Concen: 57.860 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY011076.D
Acq: 24 Oct 2022 17:52




Instrument :
MSVOA_Y
ClientSampleId :
CHRT-20430


Tgt Ion: 98 Resp: 568151
Ion Ratio Lower Upper
98 100
100 64.8 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 52.938 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.006 min
Lab File: VY011076.D
Acq: 24 Oct 2022 17:52

Tgt Ion: 95 Resp: 183507
Ion Ratio Lower Upper
95 100
174 82.6 0.0 177.0
176 79.9 0.0 175.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.489 min Scan# 11

Delta R.T. -0.001 min

Lab File: VY011076.D

Acq: 24 Oct 2022 17:52

Instrument :

MSVOA_Y

ClientSampleId :

CHRT-20430





Tgt Ion:117 Resp: 398274

Ion Ratio Lower Upper

117 100

82 54.0 43.0 64.4

119 32.1 25.4 38.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. 0.000 min

Lab File: VY011076.D

Acq: 24 Oct 2022 17:52

Tgt Ion:152 Resp: 172418

Ion Ratio Lower Upper

152 100

115 58.7 28.2 84.7

150 157.0 0.0 347.6

