

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102422\
 Data File : VY011079.D
 Acq On : 24 Oct 2022 19:00
 Operator : KP/MD
 Sample : N5256-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-F1-14.5-15.0

Quant Time: Oct 25 05:48:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	124274	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	214470	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	198759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	89049	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	59533	52.005	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.020%
35) Dibromofluoromethane	7.716	113	58479	48.867	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.740%
50) Toluene-d8	10.179	98	173710	38.791	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	77.580%
62) 4-Bromofluorobenzene	12.477	95	61318	37.912	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.820%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102422\
Data File : VY011079.D
Acq On : 24 Oct 2022 19:00
Operator : KP/MD
Sample : N5256-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-F1-14.5-15.0

Quant Time: Oct 25 05:48:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY011079.D
Acq: 24 Oct 2022 19:00




Instrument :
MSVOA_Y
ClientSampleId :
SB-F1-14.5-15.0


Tgt Ion:168 Resp: 124274
Ion Ratio Lower Upper
168 100
99 54.0 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 52.005 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY011079.D
Acq: 24 Oct 2022 19:00

Tgt Ion: 65 Resp: 59533
Ion Ratio Lower Upper
65 100
67 55.7 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY011079.D

Acq: 24 Oct 2022 19:00

Instrument :

MSVOA_Y

ClientSampleId :

SB-F1-14.5-15.0





Tgt Ion:114 Resp: 214470

Ion Ratio Lower Upper

114 100

63 19.3 0.0 43.2

88 15.7 0.0 31.8



#35

Dibromofluoromethane

Concen: 48.867 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.006 min

Lab File: VY011079.D

Acq: 24 Oct 2022 19:00




Tgt Ion:113 Resp: 58479


Ion Ratio Lower Upper

113 100

111 103.0 82.6 124.0

192 19.5 17.6 26.4









#50
Toluene-d8
Concen: 38.791 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY011079.D
Acq: 24 Oct 2022 19:00




Instrument :
MSVOA_Y
ClientSampleId :
SB-F1-14.5-15.0


Tgt Ion: 98 Resp: 173710
Ion Ratio Lower Upper
98 100
100 64.6 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 37.912 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.006 min
Lab File: VY011079.D
Acq: 24 Oct 2022 19:00

Tgt Ion: 95 Resp: 61318
Ion Ratio Lower Upper
95 100
174 85.8 0.0 177.0
176 81.2 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY011079.D
Acq: 24 Oct 2022 19:00

Instrument :
MSVOA_Y
ClientSampleId :
SB-F1-14.5-15.0

Tgt Ion:117 Resp: 198759
Ion Ratio Lower Upper
117 100
82 55.0 43.0 64.4
119 32.4 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. 0.000 min
Lab File: VY011079.D
Acq: 24 Oct 2022 19:00

Tgt Ion:152 Resp: 89049
Ion Ratio Lower Upper
152 100
115 57.9 28.2 84.7
150 158.5 0.0 347.6

