

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
 Data File : VY023564.D
 Acq On : 24 Oct 2025 11:21
 Operator : SY/MD
 Sample : Q3413-01
 Misc : 5.21g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BP-VPB-184-GW-980-982

Quant Time: Oct 24 23:49:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	812743	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	1310630	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1133524	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	486654	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	369964	54.436	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.880%
35) Dibromofluoromethane	7.640	113	418712	51.993	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.980%
50) Toluene-d8	10.109	98	1455594	48.535	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.060%
62) 4-Bromofluorobenzene	12.407	95	455763	46.030	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102425\
Data File : VY023564.D
Acq On : 24 Oct 2025 11:21
Operator : SY/MD
Sample : Q3413-01
Misc : 5.21g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BP-VPB-184-GW-980-982

Quant Time: Oct 24 23:49:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

