

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102524\
 Data File : VY020023.D
 Acq On : 25 Oct 2024 13:49
 Operator : SY/MD
 Sample : P4509-02
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AU-06-10232024

Quant Time: Oct 26 00:21:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	4113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	5652	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	2599	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	388	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	2072	40.914	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.820%
35) Dibromofluoromethane	7.634	113	1860	49.870	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.740%
50) Toluene-d8	10.103	98	5322	38.640	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	77.280%
62) 4-Bromofluorobenzene	12.401	95	540	10.851	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	21.700%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102524\
Data File : VY020023.D
Acq On : 25 Oct 2024 13:49
Operator : SY/MD
Sample : P4509-02
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AU-06-10232024

Quant Time: Oct 26 00:21:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

