

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102622\
 Data File : VY011114.D
 Acq On : 26 Oct 2022 14:27
 Operator : KP/MD
 Sample : N5281-03
 Misc : 5.68g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-3

Quant Time: Oct 27 04:38:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	266603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	470008	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	415828	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	184849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	109814	44.716	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.440%
35) Dibromofluoromethane	7.710	113	119542	45.431	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.860%
50) Toluene-d8	10.179	98	387861	39.522	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	79.040%
62) 4-Bromofluorobenzene	12.477	95	137236	38.789	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102622\
Data File : VY011114.D
Acq On : 26 Oct 2022 14:27
Operator : KP/MD
Sample : N5281-03
Misc : 5.68g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-3

Quant Time: Oct 27 04:38:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

