

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102622\
 Data File : VY011123.D
 Acq On : 26 Oct 2022 17:51
 Operator : KP/MD
 Sample : N5281-12
 Misc : 4.86g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-11

Quant Time: Oct 27 04:40:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	322680	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	570637	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	506238	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	230060	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	121402	40.844	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.680%
35) Dibromofluoromethane	7.710	113	130956	40.772	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.540%
50) Toluene-d8	10.173	98	424635	35.639	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	71.280%
62) 4-Bromofluorobenzene	12.477	95	153478	35.471	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	70.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102622\
Data File : VY011123.D
Acq On : 26 Oct 2022 17:51
Operator : KP/MD
Sample : N5281-12
Misc : 4.86g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-11

Quant Time: Oct 27 04:40:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

