

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102622\
 Data File : VY011127.D
 Acq On : 26 Oct 2022 19:21
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 27 04:41:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	74	0.00
2 T	Dichlorodifluoromethane	0.332	0.278	16.3	65	0.00
3 P	Chloromethane	0.476	0.383	19.5	63	0.00
4 C	Vinyl Chloride	0.539	0.446	17.3#	63	0.00
5 T	Bromomethane	0.373	0.309	17.2	68	0.00
6 T	Chloroethane	0.360	0.321	10.8	68	0.00
7 T	Trichlorofluoromethane	0.789	0.732	7.2	71	0.00
8 T	Diethyl Ether	0.295	0.292	1.0	74	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.495	0.478	3.4	73	0.00
10 T	Methyl Iodide	0.606	0.531	12.4	63	0.00
11 T	Tert butyl alcohol	0.062	0.067	-8.1	101	0.00
12 CM	1,1-Dichloroethene	0.474	0.434	8.4#	69	0.00
13 T	Acrolein	0.074	0.062	16.2	62	0.00
14 T	Allyl chloride	0.752	0.699	7.0	69	0.00
15 T	Acrylonitrile	0.154	0.174	-13.0	85	0.00
16 T	Acetone	0.134	0.123	8.2	69	0.00
17 T	Carbon Disulfide	1.391	1.027	26.2#	58	0.00
18 T	Methyl Acetate	0.428	0.413	3.5	83	0.00
19 T	Methyl tert-butyl Ether	1.325	1.417	-6.9	78	0.00
20 T	Methylene Chloride	0.761	0.615	19.2	79	0.00
21 T	trans-1,2-Dichloroethene	0.535	0.496	7.3	70	0.00
22 T	Diisopropyl ether	1.611	1.718	-6.6	77	0.00
23 T	Vinyl Acetate	0.963	1.058	-9.9	77	0.00
24 P	1,1-Dichloroethane	0.973	0.992	-2.0	76	0.00
25 T	2-Butanone	0.203	0.213	-4.9	79	0.00
26 T	2,2-Dichloropropane	0.864	0.806	6.7	71	0.00
27 T	cis-1,2-Dichloroethene	0.617	0.627	-1.6	75	0.00
28 T	Bromochloromethane	0.303	0.389	-28.4#	77	0.00
29 T	Tetrahydrofuran	0.125	0.139	-11.2	82	0.00
30 C	Chloroform	0.980	1.042	-6.3#	78	0.00
31 T	Cyclohexane	0.881	0.726	17.6	65	0.00
32 T	1,1,1-Trichloroethane	0.862	0.872	-1.2	75	0.00
33 S	1,2-Dichloroethane-d4	0.461	0.409	11.3	68	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	75	0.00
35 S	Dibromofluoromethane	0.304	0.257	15.5	69	0.00
36 T	1,1-Dichloropropene	0.465	0.430	7.5	70	0.00
37 T	Ethyl Acetate	0.258	0.289	-12.0	83	0.00
38 T	Carbon Tetrachloride	0.480	0.472	1.7	74	0.00
39 T	Methylcyclohexane	0.557	0.489	12.2	66	0.00
40 TM	Benzene	1.366	1.333	2.4	74	0.00
41 T	Methacrylonitrile	0.140	0.147	-5.0	87	0.00
42 TM	1,2-Dichloroethane	0.378	0.393	-4.0	78	0.00
43 T	Isopropyl Acetate	0.460	0.498	-8.3	80	0.00
44 TM	Trichloroethene	0.370	0.356	3.8	73	0.00
45 C	1,2-Dichloropropane	0.347	0.364	-4.9#	78	0.00
46 T	Dibromomethane	0.193	0.207	-7.3	80	0.00
47 T	Bromodichloromethane	0.466	0.504	-8.2	80	0.00
48 T	Methyl methacrylate	0.206	0.225	-9.2	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102622\
 Data File : VY011127.D
 Acq On : 26 Oct 2022 19:21
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 27 04:41:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	84	0.00
50 S	Toluene-d8	1.044	0.860	17.6	62	0.00
51 T	4-Methyl-2-Pentanone	0.257	0.297	-15.6	85	0.00
52 CM	Toluene	0.848	0.849	-0.1#	74	0.00
53 T	t-1,3-Dichloropropene	0.484	0.502	-3.7	77	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.563	-1.6	75	0.00
55 T	1,1,2-Trichloroethane	0.279	0.308	-10.4	82	0.00
56 T	Ethyl methacrylate	0.370	0.418	-13.0	81	0.00
57 T	1,3-Dichloropropane	0.479	0.512	-6.9	79	0.00
58 T	2-Chloroethyl Vinyl ether	0.164	0.202	-23.2	78	0.00
59 T	2-Hexanone	0.176	0.205	-16.5	84	0.00
60 T	Dibromochloromethane	0.332	0.369	-11.1	83	0.00
61 T	1,2-Dibromoethane	0.267	0.283	-6.0	80	0.00
62 S	4-Bromofluorobenzene	0.418	0.336	19.6	69	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.389	0.377	3.1	75	0.00
65 PM	Chlorobenzene	1.027	1.016	1.1	76	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.398	-4.7	81	0.00
67 C	Ethyl Benzene	1.816	1.802	0.8#	75	0.00
68 T	m/p-Xylenes	0.699	0.684	2.1	75	0.00
69 T	o-Xylene	0.663	0.670	-1.1	76	0.00
70 T	Styrene	1.124	1.171	-4.2	78	0.00
71 P	Bromoform	0.234	0.259	-10.7	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	3.633	3.416	6.0	76	0.00
74 T	N-amyl acetate	0.922	0.927	-0.5	80	0.00
75 P	1,1,2,2-Tetrachloroethane	0.762	0.778	-2.1	85	0.00
76 T	1,2,3-Trichloropropane	0.568	0.543	4.4	80	0.00
77 T	Bromobenzene	0.849	0.809	4.7	78	0.00
78 T	n-propylbenzene	4.420	4.172	5.6	76	0.00
79 T	2-Chlorotoluene	2.489	2.374	4.6	78	0.00
80 T	1,3,5-Trimethylbenzene	3.031	2.876	5.1	76	0.00
81 T	trans-1,4-Dichloro-2-butene	0.264	0.265	-0.4	80	0.00
82 T	4-Chlorotoluene	2.592	2.430	6.2	77	0.00
83 T	tert-Butylbenzene	2.621	2.573	1.8	78	0.00
84 T	1,2,4-Trimethylbenzene	2.989	2.871	3.9	77	0.00
85 T	sec-Butylbenzene	3.938	3.787	3.8	77	0.00
86 T	p-Isopropyltoluene	3.267	3.131	4.2	76	0.00
87 T	1,3-Dichlorobenzene	1.695	1.612	4.9	78	0.00
88 T	1,4-Dichlorobenzene	1.685	1.597	5.2	78	0.00
89 T	n-Butylbenzene	3.053	2.908	4.7	75	0.00
90 T	Hexachloroethane	0.645	0.632	2.0	80	0.00
91 T	1,2-Dichlorobenzene	1.516	1.473	2.8	80	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.126	0.132	-4.8	88	0.00
93 T	1,2,4-Trichlorobenzene	0.906	0.861	5.0	76	0.00
94 T	Hexachlorobutadiene	0.533	0.512	3.9	79	0.00
95 T	Naphthalene	1.845	1.927	-4.4	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102622\
Data File : VY011127.D
Acq On : 26 Oct 2022 19:21
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 27 04:41:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.791	0.778	1.6	79	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6