

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102722\
 Data File : VY011130.D
 Acq On : 27 Oct 2022 11:12
 Operator : KP/MD
 Sample : VY1027SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1027SBL01

Quant Time: Oct 28 00:45:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	196270	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	337107	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	302621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	138878	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	93607	51.776	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.560%
35) Dibromofluoromethane	7.716	113	97659	52.059	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.120%
50) Toluene-d8	10.179	98	306138	43.493	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	86.980%
62) 4-Bromofluorobenzene	12.477	95	108642	43.152	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102722\
Data File : VY011130.D
Acq On : 27 Oct 2022 11:12
Operator : KP/MD
Sample : VY1027SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1027SBL01

Quant Time: Oct 28 00:45:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

