

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102722\
 Data File : VY011133.D
 Acq On : 27 Oct 2022 12:42
 Operator : KP/MD
 Sample : N5265-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11976

Quant Time: Oct 27 12:58:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	288923	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	497574	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	425501	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	168905	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	119191	44.785	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.580%
35) Dibromofluoromethane	7.716	113	126553	45.431	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.860%
50) Toluene-d8	10.179	98	386706	37.221	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	74.440%
62) 4-Bromofluorobenzene	12.477	95	132458	35.075	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	70.140%
Target Compounds						
20) Methylene Chloride	4.710	84	66062	15.220	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102722\
Data File : VY011133.D
Acq On : 27 Oct 2022 12:42
Operator : KP/MD
Sample : N5265-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11976

Quant Time: Oct 27 12:58:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

