

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102722\
 Data File : VY011138.D
 Acq On : 27 Oct 2022 14:56
 Operator : KP/MD
 Sample : N5281-05
 Misc : 6.23g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-5S

Quant Time: Oct 28 00:47:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	246411	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	433669	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	389181	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	174897	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	121092	53.349	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.700%	
35) Dibromofluoromethane	7.710	113	134691	55.975	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.960%	
50) Toluene-d8	10.179	98	424182	46.845	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 93.680%	
62) 4-Bromofluorobenzene	12.477	95	155583	48.407	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 96.820%	
Target Compounds						
20) Methylene Chloride	4.710	84	41303	9.227	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102722\
Data File : VY011138.D
Acq On : 27 Oct 2022 14:56
Operator : KP/MD
Sample : N5281-05
Misc : 6.23g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-5S

Quant Time: Oct 28 00:47:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

