

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102725\
 Data File : VY023592.D
 Acq On : 27 Oct 2025 12:30
 Operator : SY/MD
 Sample : Q3466-04
 Misc : 4.79g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B9-0.5-1.0-20251023

Quant Time: Oct 28 04:39:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	745194	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1231218	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	979970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	361763	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	314555	50.479	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.960%
35) Dibromofluoromethane	7.634	113	380851	50.342	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.680%
50) Toluene-d8	10.109	98	1364092	48.417	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.840%
62) 4-Bromofluorobenzene	12.408	95	359173	38.614	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102725\
Data File : VY023592.D
Acq On : 27 Oct 2025 12:30
Operator : SY/MD
Sample : Q3466-04
Misc : 4.79g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B9-0.5-1.0-20251023

Quant Time: Oct 28 04:39:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

