

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102725\
 Data File : VY023594.D
 Acq On : 27 Oct 2025 13:17
 Operator : SY/MD
 Sample : Q3469-03
 Misc : 5.29g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-13-VOC

Quant Time: Oct 28 04:39:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	737825	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	1237706	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1039194	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	419726	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	326935	52.990	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.980%
35) Dibromofluoromethane	7.640	113	385240	50.655	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.320%
50) Toluene-d8	10.109	98	1393294	49.194	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.380%
62) 4-Bromofluorobenzene	12.401	95	388307	41.528	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102725\
Data File : VY023594.D
Acq On : 27 Oct 2025 13:17
Operator : SY/MD
Sample : Q3469-03
Misc : 5.29g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-13-VOC

Quant Time: Oct 28 04:39:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

