

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102725\
 Data File : VY023596.D
 Acq On : 27 Oct 2025 15:47
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 28 04:40:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.360	0.276	23.3	70	0.00
3 P	Chloromethane	0.490	0.390	20.4	70	0.00
4 C	Vinyl Chloride	0.639	0.528	17.4#	72	0.00
5 T	Bromomethane	0.536	0.443	17.4	76	0.01
6 T	Chloroethane	0.435	0.390	10.3	76	0.00
7 T	Trichlorofluoromethane	0.887	0.771	13.1	74	0.00
8 T	Diethyl Ether	0.232	0.205	11.6	77	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.460	0.435	5.4	81	0.00
10 T	Methyl Iodide	0.512	0.538	-5.1	81	0.00
11 T	Tert butyl alcohol	0.031	0.021	32.3#	67	0.00
12 CM	1,1-Dichloroethene	0.446	0.408	8.5#	77	0.00
13 T	Acrolein	0.044	0.009	79.5#	17#	0.00
14 T	Allyl chloride	0.544	0.501	7.9	77	0.00
15 T	Acrylonitrile	0.089	0.079	11.2	79	0.00
16 T	Acetone	0.100	0.079	21.0	76	0.00
17 T	Carbon Disulfide	1.277	1.087	14.9	72	0.00
18 T	Methyl Acetate	0.261	0.252	3.4	83	0.00
19 T	Methyl tert-butyl Ether	1.072	0.993	7.4	79	0.00
20 T	Methylene Chloride	0.538	0.473	12.1	83	0.00
21 T	trans-1,2-Dichloroethene	0.491	0.478	2.6	82	0.01
22 T	Diisopropyl ether	1.202	1.174	2.3	82	0.00
23 T	Vinyl Acetate	0.651	0.586	10.0	76	0.00
24 P	1,1-Dichloroethane	0.796	0.780	2.0	83	0.00
25 T	2-Butanone	0.122	0.097	20.5	75	0.00
26 T	2,2-Dichloropropane	0.731	0.709	3.0	81	0.00
27 T	cis-1,2-Dichloroethene	0.567	0.566	0.2	85	0.00
28 T	Bromochloromethane	0.297	0.267	10.1	81	0.00
29 T	Tetrahydrofuran	0.069	0.057	17.4	72	0.00
30 C	Chloroform	0.875	0.885	-1.1#	86	0.00
31 T	Cyclohexane	0.696	0.605	13.1	76	0.00
32 T	1,1,1-Trichloroethane	0.795	0.779	2.0	82	0.00
33 S	1,2-Dichloroethane-d4	0.418	0.388	7.2	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.307	0.307	0.0	83	0.00
36 T	1,1-Dichloropropene	0.407	0.400	1.7	82	0.00
37 T	Ethyl Acetate	0.161	0.144	10.6	78	0.00
38 T	Carbon Tetrachloride	0.493	0.482	2.2	81	0.00
39 T	Methylcyclohexane	0.530	0.505	4.7	78	0.00
40 TM	Benzene	1.288	1.295	-0.5	84	0.00
41 T	Methacrylonitrile	0.087	0.080	8.0	81	0.00
42 TM	1,2-Dichloroethane	0.332	0.320	3.6	83	0.00
43 T	Isopropyl Acetate	0.321	0.282	12.1	76	0.00
44 TM	Trichloroethene	0.377	0.374	0.8	83	0.00
45 C	1,2-Dichloropropane	0.285	0.287	-0.7#	85	0.00
46 T	Dibromomethane	0.180	0.178	1.1	83	0.00
47 T	Bromodichloromethane	0.453	0.457	-0.9	85	0.00
48 T	Methyl methacrylate	0.151	0.132	12.6	73	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102725\
 Data File : VY023596.D
 Acq On : 27 Oct 2025 15:47
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 28 04:40:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	75	0.00
50 S	Toluene-d8	1.144	1.144	0.0	82	0.00
51 T	4-Methyl-2-Pentanone	0.166	0.150	9.6	77	0.00
52 CM	Toluene	0.836	0.858	-2.6#	84	0.00
53 T	t-1,3-Dichloropropene	0.394	0.380	3.6	80	0.00
54 T	cis-1,3-Dichloropropene	0.465	0.461	0.9	82	0.00
55 T	1,1,2-Trichloroethane	0.242	0.236	2.5	83	0.00
56 T	Ethyl methacrylate	0.279	0.272	2.5	79	0.00
57 T	1,3-Dichloropropane	0.383	0.377	1.6	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.131	-4.8	82	0.00
59 T	2-Hexanone	0.119	0.102	14.3	74	0.00
60 T	Dibromochloromethane	0.339	0.338	0.3	84	0.00
61 T	1,2-Dibromoethane	0.231	0.220	4.8	80	0.00
62 S	4-Bromofluorobenzene	0.378	0.379	-0.3	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.516	0.522	-1.2	82	0.00
65 PM	Chlorobenzene	1.087	1.083	0.4	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.388	0.389	-0.3	86	0.00
67 C	Ethyl Benzene	1.760	1.799	-2.2#	84	0.00
68 T	m/p-Xylenes	0.706	0.725	-2.7	83	0.00
69 T	o-Xylene	0.662	0.678	-2.4	83	0.00
70 T	Styrene	1.089	1.122	-3.0	84	0.00
71 P	Bromoform	0.240	0.226	5.8	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.295	3.346	-1.5	85	0.00
74 T	N-amyl acetate	0.592	0.501	15.4	73	0.00
75 P	1,1,2,2-Tetrachloroethane	0.539	0.482	10.6	82	0.00
76 T	1,2,3-Trichloropropane	0.441	0.362	17.9	71	0.00
77 T	Bromobenzene	0.876	0.836	4.6	84	0.00
78 T	n-propylbenzene	3.813	3.858	-1.2	84	0.00
79 T	2-Chlorotoluene	2.233	2.208	1.1	84	0.00
80 T	1,3,5-Trimethylbenzene	2.663	2.724	-2.3	85	0.00
81 T	trans-1,4-Dichloro-2-butene	0.180	0.152	15.6	76	0.00
82 T	4-Chlorotoluene	2.300	2.281	0.8	85	0.00
83 T	tert-Butylbenzene	2.458	2.492	-1.4	84	0.00
84 T	1,2,4-Trimethylbenzene	2.657	2.709	-2.0	85	0.00
85 T	sec-Butylbenzene	3.527	3.596	-2.0	85	0.00
86 T	p-Isopropyltoluene	3.003	3.074	-2.4	84	0.00
87 T	1,3-Dichlorobenzene	1.706	1.638	4.0	84	0.00
88 T	1,4-Dichlorobenzene	1.685	1.589	5.7	83	0.00
89 T	n-Butylbenzene	2.616	2.639	-0.9	83	0.00
90 T	Hexachloroethane	0.646	0.635	1.7	86	0.00
91 T	1,2-Dichlorobenzene	1.496	1.434	4.1	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.088	0.074	15.9	75	0.00
93 T	1,2,4-Trichlorobenzene	0.932	0.862	7.5	80	0.00
94 T	Hexachlorobutadiene	0.581	0.555	4.5	81	0.00
95 T	Naphthalene	1.485	1.282	13.7	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102725\
Data File : VY023596.D
Acq On : 27 Oct 2025 15:47
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 28 04:40:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100725S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.808	0.733	9.3	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6