

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
 Data File : VY011152.D
 Acq On : 28 Oct 2022 12:24
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Oct 28 23:44:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	276762	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	476166	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	420219	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	187903	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	152590	53.320	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.640%
35) Dibromofluoromethane	7.716	113	149455	51.039	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.080%
50) Toluene-d8	10.179	98	571897	49.298	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.600%
62) 4-Bromofluorobenzene	12.477	95	182165	47.991	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102822\
Data File : VY011152.D
Acq On : 28 Oct 2022 12:24
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Oct 28 23:44:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

